

Freeze-Thaw Stability Test Result

[27900 Human Titin N-Fragment \(Urine\) ELISA Kit - IBL](#)

Sample Type: **Human Urine**

| Sample | | Result (pmol/L) | Dilution | Calc. (nmol/L) | % |
|--------|-----|-----------------|----------|----------------|------|
| No.1 | FT0 | 1350.7 | 10 | 13.51 | 100 |
| | FT1 | 1191.9 | 10 | 11.92 | 88.2 |
| | FT2 | 1247.9 | 10 | 12.48 | 92.4 |
| | FT3 | 1214.2 | 10 | 12.14 | 89.9 |
| No.2 | FT0 | 398.4 | 10 | 3.98 | 100 |
| | FT1 | 356.1 | 10 | 3.56 | 89.4 |
| | FT2 | 361.1 | 10 | 3.61 | 90.6 |
| | FT3 | 371.9 | 10 | 3.72 | 93.3 |
| No.3 | FT0 | 293.8 | 10 | 2.94 | 100 |
| | FT1 | 269.5 | 10 | 2.70 | 91.7 |
| | FT2 | 274 | 10 | 2.74 | 93.3 |
| | FT3 | 275.9 | 10 | 2.76 | 93.9 |

It is stable up to 3 times freeze-thaw repeat.

