

Product information



User's Manual

Mouse BDNF ELISA

For the precise analytical measurement of BDNF in serum, plasma, body fluids, tissue lysates or cell culture supernates.

BE69100



96

Storage:

2–8 °C

RUO

For Research Use Only – Not for Use in Diagnostic Procedures.

Manufactured for:

Immuno-Biological Laboratories, Inc. (IBL-America)

8201 Central Ave NE, Suite P, Minneapolis, MN 55432

Toll Free: (888) 523-1246 Fax: (763) 780-2988

www.ibl-america.com / info@ibl-america.com

1 INTRODUCTION

1.1 Intended Use

The IBL-America mouse BDNF ELISA has been designed for the precise analytical measurement of BDNF concentrations in serum, plasma, body fluids, tissue lysates or cell culture supernates from mice. For research use only, not for use in diagnostic procedures.

1.2 Background

Brain-derived neurotrophic factor, also known as BDNF, is a prosurvival factor induced by cortical neurons that is necessary for survival of striatal neurons in the brain. BDNF is a member of the "neurotrophin" family of growth factors, which are related to the NGF. It acts on certain neurons of the central nervous system and the peripheral nervous system, helping to support the survival of existing neurons, and encourage the growth and differentiation of new neurons and synapses. In the brain, it is active in the hippocampus, cortex, and basal forebrain—areas vital to learning, memory, and higher thinking. Post mortem analysis has shown lowered levels of BDNF in the brain tissues of people with Alzheimer's disease.

2 PRINCIPLE OF THE TEST

This kit is based on a sandwich enzyme-linked immune-sorbent assay technology. Anti-BDNF polyclonal antibody was pre-coated onto 96-well plates. The biotin conjugated anti-BDNF polyclonal second antibody was used as detection antibody. The standards, test samples and biotin conjugated detection antibody are added to the wells subsequently, and wash with wash buffer. Avidin-Biotin-Peroxidase Complex was added and unbound conjugates were washed away with wash buffer. TMB substrate is added to visualize HRP enzymatic reaction. TMB is catalyzed by HRP to produce a blue color product that changes into yellow after adding acidic stop solution. The optical density of the yellow color is proportional to the BDNF amount of the sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of BDNF in the sample can be calculated.

3 WARNINGS AND PRECAUTIONS

1. This kit is for research use only, not for use in diagnostic procedures.
2. Before the experiment, centrifuge each kit component for several minutes to bring down all reagents to the bottom of tubes.
3. It is recommend to measure each standard and sample in duplicate.
4. Do NOT let the plate completely dry at any time! Since the dry condition can inactivate the biological material on the plate.
5. Do not reuse pipette tips and tubes to avoid cross contamination.
6. Do not use the expired components or the components from different lot numbers.
7. To avoid the marginal effect of plate incubation for temperature differences (the marginal wells always get stronger reaction), it is recommend to equilibrate the ABC working solution and TMB substrate for at least 30 min at 37°C before adding to wells.
8. The TMB substrate (Kit Component 8) is colorless and transparent before use. If not, please contact us for replacement.

4 REAGENTS

4.1 Reagents provided

1. One 96-well microtiterplate pre-coated with anti-mouse BDNF antibody
2. Lyophilized mouse BDNF standards: 2 tubes (10 ng / tube)
3. Sample / Standard diluent buffer: 30 ml
4. Biotin conjugated anti-Mouse BDNF antibody (Concentrated): 130µl. Dilution: 1:100
5. Antibody diluent buffer: 12 ml
6. Avidin-Biotin-Peroxidase Complex (ABC) (Concentrated): 130µl. Dilution: 1:100
7. ABC diluent buffer: 12 ml
8. TMB substrate: 10 ml
9. Stop solution: 10 ml
10. Wash buffer (25X): 30 ml

4.2 Materials required but not provided

1. 37°C incubator
2. Microtiterplate reader (wavelength: 450 nm)
3. Precise pipette and disposable pipette tips
4. Automated plate washer
5. ELISA shaker
6. 1.5 ml Eppendorf tubes
7. Plate cover
8. Absorbent filter papers
9. Plastic or glass container with volume of above 1 L

4.3 Storage Conditions / Expiration

Store at 2-8°C for 6 months, at -20°C for 12 months.

4.4 Preparation of sample and reagents

1. Sample

Isolate the test samples soon after collecting, then, analyze immediately (within 2 hours). Or aliquot and store at -20°C for long term. Avoid multiple freeze-thaw cycles.

◇ **Body fluids, tissue lysates and cell culture supernatants:** Centrifuge to remove precipitate, analyze immediately or aliquot and store at -20°C.

◇ **Serum:** Coagulate the serum at room temperature (about 4 hours). Centrifuge at approximately $2000 \times g$ for 20 min. Analyze the serum immediately or aliquot and store at -20°C

◇ **Plasma:** Collect plasma with citrate, heparin or EDTA as the anticoagulant. Centrifuge for 15min at 2-8°C at $1000 \times g$ within 30 min of collection. For eliminating the platelet effect, suggesting that further centrifugation for 10 min at 2-8°C at $10000 \times g$. Analyze immediately or aliquot and store frozen at -20°C.

Note: 1. Coagulate blood samples completely, then centrifuge, and avoid hemolysis and particles.

2. NaN_3 cannot be used as test sample preservative, since it is the inhibitor for HRP.

>> Sample Dilution Guideline

End user should estimate the concentration of the target protein in the test sample first, and select a proper dilution factor to make the diluted target protein concentration falls the optimal detection range of the kit. Dilute the sample with the provided diluent buffer, and several trials may be necessary in practice. The test sample must be well mixed with the diluent buffer.

◇ **High target protein concentration (20-200 ng/ml):** Dilution: 1:100. i.e. Add 1 μl of sample into 99 μl of Sample / Standard diluent buffer (Kit Component 3).

◇ **Medium target protein concentration (2-20 ng/ml):** Dilution: 1:10. i.e. Add 10 μl of sample into 90 μl of Sample / Standard diluent buffer (Kit Component 3).

◇ **Low target protein concentration (31.2-2000 pg/ml):** Dilution: 1:2. i.e. Add 50 μl of sample into 50 μl of Sample / Standard diluent buffer (Kit Component 3).

◇ **Very low target protein concentration ($\leq 31.2 \text{ pg/ml}$):** Unnecessary to dilute, or dilute at 1:2.

2. Wash buffer

Dilute the concentrated Wash buffer 25-fold (1:25) with distilled water (i.e. add 30 ml of concentrated wash buffer into 720 ml of distilled water).


3. Standard

Reconstitution of the lyophilized mouse BDNF standard (Kit Component 2): standard solution should be prepared no more than 2 hours prior to the experiment. Two tubes of standard are included in each kit. Use one tube for each experiment. (**Note: Do not dilute the standard directly in the plate**)

a. 10,000 pg/ml of standard solution: Add **1 mL** of Sample / Standard diluent buffer (Kit Component 3) into one Standard (Kit Component 2) tube, keep the tube at room temperature for 10 min and mix thoroughly.

b. 2000 pg/ml of standard solution: Add **0.2 mL** of the above 10 ng/ml standard solution into **0.8 mL** sample diluent buffer (Kit Component 3) and mix thoroughly.

c. 1000 pg/ml → 31.2 pg/ml of standard solutions: Label 6 Eppendorf tubes with 1000 pg/ml, 500 pg/ml, 250 pg/ml, 125 pg/ml, 62.5 pg/ml, 31.2 pg/ml, respectively. Aliquot **0.3 mL** of the Sample / Standard diluent buffer (Kit Component 3) into each tube. Add **0.3 mL** of the above 2000 pg/ml standard solution into 1st tube and mix thoroughly. Transfer **0.3 mL** from 1st tube to 2nd tube and mix thoroughly. Transfer **0.3 mL** from 2nd tube to 3rd tube and mix thoroughly, and so on.



Note: The standard solutions are best used within 2 hours. The 10,000 pg/ml standard solution should be used within 12 hours. Or store at -20°C for up to 48 hours. Avoid repeated freeze-thaw cycles.

4. Preparation of Biotin conjugated anti-mouse BDNF antibody (Kit Component 4) working solution: prepare no more than 2 hours before the experiment.

a. Calculate the total volume of the working solution: 0.1 ml / well \times quantity of wells. (Allow 0.1-0.2 ml more than the total volume)

b. Dilute the Biotin conjugated anti-Mouse BDNF antibody (Kit Component 4) with Antibody diluent buffer (Kit Component 5) at 1:100 and mix thoroughly. i.e. Add 1 μl of Biotin conjugated anti-Mouse BDNF antibody into 99 μl of Antibody diluent buffer.

- 5. Preparation of Avidin-Biotin-Peroxidase Complex (ABC) (Kit Component 6) working solution:** prepare no more than 1 hour before the experiment.
- Calculate the total volume of the working solution: 0.1 ml / well \times quantity of wells. (Allow 0.1-0.2 ml more than the total volume)
 - Dilute the Avidin-Biotin-Peroxidase Complex (ABC) (Kit Component 6) with ABC diluent buffer (Kit Component 7) at 1:100 and mix thoroughly. i.e. Add 1 μ l of Avidin-Biotin-Peroxidase Complex (ABC) into 99 μ l of ABC diluent buffer.

5 ASSAY PROCEDURE

5.1 General Remarks

Before adding to wells, equilibrate the ABC working solution and TMB substrate (Kit Component 8) for at least 30 minutes at 37°C. It is recommended to plot a standard curve for each test.

5.2 Test Procedure

- Set standard, test sample and control (zero) wells on the pre-coated plate respectively, and then, record their positions. It is recommended to measure each standard and sample in duplicate.
- Aliquot 0.1ml of 2000 pg/ml, 1000 pg/ml, 500 pg/ml, 250 pg/ml, 125 pg/ml, 62.5 pg/ml, 31.2 pg/ml standard solutions into the standard wells.
- Add 0.1 ml of Sample / Standard diluent buffer (Kit Component 3) into the control (zero) well.
- Add 0.1 ml of properly diluted sample (Mouse serum, plasma, body fluids, tissue lysates or cell culture supernatants) into test sample wells.
- Seal the plate with a cover and incubate at 37°C for 90 min.
- Remove the cover and discard the plate content, clap the plate on the absorbent filter papers or other absorbent material. **Do NOT let the wells completely dry at any time. Do not wash plate!**
- Add 0.1 ml of Biotin conjugated anti-Mouse BDNF antibody work solution into the above wells (standard, test sample & zero wells). Add the solution at the bottom of each well without touching the side wall.
- Seal the plate with a cover and incubate at 37°C for 60 min.
- Remove the cover, and wash plate 3 times with Wash buffer (Kit Component 10) using one of the following methods:

Manual Washing: Discard the solution in the plate without touching the side walls. Clap the plate on absorbent filter papers or other absorbent material. Fill each well completely with Wash buffer (Kit Component 10) buffer and vortex mildly on ELISA shaker for 2 min, then aspirate contents from the plate, and clap the plate on absorbent filter papers or other absorbent material. Repeat this procedure two more times for a **total of THREE washes**.

- Automated Washing:** Aspirate all wells, then wash plate **THREE times** with Wash buffer (Kit Component 10) (overfilling wells with the buffer). After the final wash, invert plate, and clap the plate on absorbent filter papers or other absorbent material. It is recommended that the washer be set for a soaking time of 1 min or shaking.
- Add 0.1 ml of ABC working solution into each well, cover the plate and incubate at 37°C for 30 min.
 - Remove the cover and wash plate 5 times with Wash buffer (Kit Component 10), and each time let the wash buffer stay in the wells for 1-2 min. (See Step 9 for plate wash method).
 - Add 0.1 ml of TMB substrate (Kit Component 8) into each well, cover the plate and incubate at 37°C in dark within 30 min. (**Note:** This incubation time is for reference use only, the optimal time should be determined by end user.) And the shades of blue can be seen in the first 3-4 wells (with most concentrated Mouse BDNF standard solutions), the other wells show no obvious color.
 - Add 0.1 ml of Stop solution (Kit Component 9) into each well and mix thoroughly. The color changes into yellow immediately.
 - Read the O.D. absorbance at 450 nm in a microplate reader within 30 min after adding the stop solution.

5.3 Results

For calculation, (the relative O.D.₄₅₀) = (the O.D.₄₅₀ of each well) – (the O.D.₄₅₀ of Zero well). The standard curve can be plotted as the relative O.D.₄₅₀ of each standard solution (Y) vs. the respective concentration of the standard solution (X). The Mouse BDNF concentration of the samples can be interpolated from the standard curve.

Note: If the samples measured were diluted, multiply the dilution factor to the concentrations from interpolation to obtain the concentration before dilution.

6 QUALITY CONTROL

Good laboratory practice requires that controls be run with each calibration curve. A statistically significant number of controls should be assayed to establish mean values and acceptable ranges to assure proper performance. It is recommended to use controls according to state and federal regulations. The use of controls is advised to assure the day to day validity of results. It is also recommended to make use of national or international Quality Assessment programs in order to ensure the accuracy of the results. Employ appropriate statistical methods for analyzing control values and trends. If the results of the assay do not fit to the established acceptable ranges of control materials, results of unknowns should be considered invalid.

In this case, please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods. After checking the above mentioned items without finding any error contact your distributor or IBL-America directly.

7 PERFORMANCE CHARACTERISTICS

7.1 Range

31.2 pg/ml - 2000 pg/ml


7.2 Sensitivity

< 2 pg/ml

7.3 Typical Data & Standard Curve

Results of a typical standard run of a Mouse BDNF ELISA Kit are shown below. **This standard curve was generated at our lab for demonstration purpose only.** Each user should obtain their own standard curve as per experiment. (N/A=not applicable)

X	pg/ml	0	31.2	62.5	125	250	500	1000	2000
Y	OD450	0.082	0.109	0.142	0.288	0.553	1.060	1.867	2.512



7.4 References:

1. Binder DK, Scharfman HE (September 2004). "Brain-derived Neurotrophic Factor". *Growth Factors* 22 (3): 123–31.
2. Maisonnier PC, Le Beau MM, Espinosa R et al. (July 1991). "Human and rat brain-derived neurotrophic factor and neurotrophin-3: gene structures, distributions, and chromosomal localizations". *Genomics* 10 (3): 558–68.
3. Maisonnier PC, Le Beau MM, Espinosa R et al. (July 1991). "Human and rat brain-derived neurotrophic factor and neurotrophin-3: gene structures, distributions, and chromosomal localizations". *Genomics* 10 (3): 558–68.
4. Yamada K, Nabeshima T (April 2003). "Brain-derived neurotrophic factor/TrkB signaling in memory processes". *J. Pharmacol. Sci.* 91 (4): 267–70.

8 ORDERING INFORMATION

This kit is manufactured for Immuno-Biological Laboratories, Inc. (IBL-America). For ordering information, please contact:

Immuno-Biological Laboratories, Inc. (IBL-America)
 8201 Central Ave NE, Suite P
 Minneapolis, MN 55432
 Toll Free: (888) 523-1246
 Fax: (763) 780-2988
www.ibl-america.com
info@ibl-america.com