

Manufactured for Immuno-Biological Laboratories Inc.
(IBL-America)
8201 Central Avenue, NE, Suite P
Minneapolis, MN 55432
Tel: 763-780-2955
Toll Free: 1-888-523-1246



West Nile Virus IgG/IgM ELISA

CONTENTS

- 1 INTENDED USE
- 2 BACKGROUND
- 3 TEST PRINCIPLE
- 4 KIT COMPONENTS
- 5 MATERIAL REQUIRED BUT NOT SUPPLIED
- 6 STORAGE AND STABILITY
- 7 TEST PROCEDURE
 - 7.1 Evidence of Deterioration
 - 7.2 Sample Preparation and Storage
 - 7.3 Preparation of Kit Reagents
 - 7.4 Overview - Test Procedure
 - 7.5 Manual Test Procedure
 - 7.6 Automated Test Procedure
 - 7.7 Positive Control / Accuracy Control
- 8 TEST EVALUATION
 - 8.1 Criteria of Validity
 - 8.2 Cut-off Calculation
 - 8.3 Borderline Ranges
- 9 PERFORMANCE CHARACTERISTICS
 - 9.1 Sensitivity and Specificity
 - 9.2 Reproducibility
- 10 SAFETY MEASURES
 - 10.1 Statements of Warning
 - 10.2 Disposal
- 11 REFERENCES

**For Research Use
Only – Not for Use
in Clinical
Procedures**

IBL-America West Nile Virus IgG/IgM ELISA

Enzyme-immunoassay for determination of human antibodies

West Nile Virus IgG ELISA

Order no.: IB05104

West Nile Virus IgM ELISA

Order no.: IB05105

For Research Use Only – Not for Use in Clinical Procedures

1 INTENDED USE

The IBL-America West Nile Virus IgG and IgM tests are qualitative immunoassays for the demonstration of human antibodies in serum or plasma directed against West Nile virus.

2 BACKGROUND

Besides the Dengue virus, the Yellow Fever virus and the TBE virus, the West Nile virus belongs to the human pathogenic species among the flaviviruses. Birds such as ravens and sparrows represent the natural reservoirs for the arbovirus (*arthropode borne virus*). Mosquitos of the genus *Culex*, *Aedes* and *Ochlerotatus* are the main vectors of the West Nile virus to humans.

ELISA test systems are especially suited for the differential analysis of immunoglobulin classes directed against the virus.

3 TEST PRINCIPLE

The ELISA (Enzyme Linked Immunosorbent Assay) is an immunoassay, which is particularly suited to the determination of antibodies in various kinds of samples. The reaction is based on the specific interaction of antibodies with their corresponding antigen. The test strips of the microtiter plate are coated with specific antigens of the pathogen of interest. If antibodies in the sample are present, they bind to the fixed antigen. A secondary antibody, which has been conjugated with the enzyme alkaline phosphatase, detects and binds to the immune complex. The colourless substrate p-nitrophenylphosphate is then converted into the coloured product p-nitrophenol. The signal intensity of this reaction product is proportional to the concentration of the analyte in the sample and is measured photometrically.

4 KIT COMPONENTS

Test Components	Pieces / Volume	
	IgG	IgM
Break apart microtiter test strips each with eight antigen coated single wells , (altogether 96) [MTP] , 1 frame. The coating material is inactivated.	12 pieces	12 pieces
Standard serum (ready-to-use) [STD] , Human serum in protein containing phosphate buffer; negative for anti-HIV Ab, HBs-Ag (Hepatitis B-Virus surface antigen) and anti-HCV Ab; preservative: < 0.1 % sodium azide; colouring: Amaranth O.	2 x 2 ml	-
Cut-off serum (ready-to-use) [C/O] Human serum in protein containing phosphate buffer, negative for anti-HIV Ab, HBs-Ag (Hepatitis B-Virus surface antigen) and anti-HCV Ab; preservative: < 0.1 % sodium azide; colouring: Chinaldin yellow	-	2 x 2 ml
Positive control serum (ready-to-use) [POS] Human serum in protein containing phosphate buffer, negative for anti-HIV Ab, HBs-Ag (Hepatitis B-Virus surface antigen) and anti-HCV Ab; preservative: < 0.1 % sodium azide; colouring: Amaranth O	-	2 ml
Negative control serum (ready-to-use) [NEG] , Human serum in protein containing phosphate buffer; negative for anti-HIV Ab, HBs-Ag (Hepatitis B-Virus surface antigen) and anti-HCV Ab; preservative: < 0.1 % sodium azide; colouring: Lissamin Green V.	2 ml	2 ml
Anti-human IgA, IgG or IgM conjugate (ready-to-use) [APC] , Anti-human IgA, IgG or IgM polyclonal antibody, conjugated to alkaline phosphatase, stabilised with protein stabilisation solution; preservative: <0.1% methylisothiazolone, <0.1% bromnitrodioxane.	13 ml	13 ml
Washing solution concentrate (sufficient for 1000 ml) [WASH] , Sodium chloride solution with Tween 20 and 30 mM Tris/HCl, pH 7.4; preservative: < 0.1 % sodium azide.	33.3 ml	33.3 ml
Dilution buffer [DILB] , Protein containing phosphate buffer with Tween 20; preservative: < 0.1 % sodium azide; colouring: 0.01 g/l Bromphenol blue.	2 x 50 ml	2 x 50 ml
Stopping solution [STOP] , <0.1 N sodium hydroxide, 40 mm EDTA	15 ml	15 ml
Substrate (ready-to-use) [pNPP] , Para-nitrophenylphosphate in solvent free buffer; preservative: < 0.1 % sodium azide	13 ml	13ml
Quality control certificate [INFO]	1 page	1 page

5 MATERIAL REQUIRED BUT NOT SUPPLIED

- common laboratory equipment
- for the IgM detection: Rf-Absorbent, order no. IB05998 (20 ml)
- photometer for microtitre plates with filter, wavelength 405 nm, recommended reference wavelength 620 nm - 690 nm (e.g. 650 nm)
- Microtiter plate washer
- incubator 37 °C
- moist chamber
- distilled water

Recommended but not required:

Control serum 5 x 3 ml order no.: IB05104CON for West Nile Virus IgG

Control serum 5 x 3 ml order no.: IB05105CON for West Nile Virus IgM

6 STORAGE AND STABILITY

Reagent	Storage	Stability
Microtiter strips (coated with antigen)	unopened after opening at 2 – 8 °C in closed aluminum bag with desiccant	see expiry date; 6 months
Control sera / Standard sera	Unopened after opening at 2 – 8 °C	see expiry date 6 months
Conjugate	Unopened after opening at 2 – 8 °C	see expiry date 6 months
Dilution buffer	Unopened after opening at 2 – 8 °C	see expiry date 6 months
Washing solution	Unopened/ after opening at 2 – 8 °C working dilution at 2 – 8 °C working dilution at room temperature	see expiry date; 2 weeks; 1 week
Substrate	Unopened after opening at 2 – 8 °C	see expiry date 6 months
Stopping solution	Unopened after opening at 2 – 8 °C	see expiry date 6 months

7 TEST PROCEDURE

7.1 Evidence of Deterioration

Optimum results can only be achieved if the instructions are strictly followed. The components of this kit must not be exchanged for reagents of other manufacturers. Standard and control sera are defined exclusively for the test kit to be used and must not be used in other lots. Dilution buffer, washing solution, substrate and stop solution can be used for all IBL-America immunoassays coded IB05xxx irrespective of the lot and the test.

Each ELISA classic test contains a ready-to-use sample dilution buffer. In some cases the use of special dilution buffers is necessary to guarantee consistent quality and reliable results. The dilution buffers can be used irrespective of the lots.

There are three different conjugate concentrations for each immunoglobulin class (IgA, IgG, IgM), indicated on the label as + (low), ++ (medium) and +++ (high). Conjugates with the same concentration and of the same immunoglobulin class are interchangeable and can be used for other SERION ELISA classic immunoassays irrespective of the lot and the test. Dilution or alteration of the reagents may result in a loss of sensitivity. Use aseptic techniques when removing aliquots from the reagent tubes to avoid contamination.

Reproducibility of test results is dependent on thorough mixing of the reagents. Agitate the flasks containing control sera before use and also all samples after dilution (e.g. by using a vortex mixer).

Be sure to pipette carefully and comply with the given incubation times and temperatures. Significant time differences between pipetting the first and last well of the microtiter plate when dispensing samples and control sera, conjugate or substrate can result in different pre-incubation times, which may influence the precision and reproducibility of the results. Avoid exposure of reagents to strong light during storage and incubation.

Adequate washing avoids test unspecificities. Therefore, the washing procedure should be carried out carefully. All of the flat bottom wells should be filled with equal volumes of washing buffer. At the end of the procedure ensure that the wells are free of all washing buffer in order to avoid uncontrolled dilution effects. Avoid foaming!

Reagents must be tightly closed after use to avoid evaporation and contamination. Take care not to mix-up the caps of the bottles and/or vials.

The ELISA immunoassay is only valid if the lot-specific validation criteria on the quality control certificate are fulfilled

7.2 Sample Preparation and Storage

Lipaemic, hemolytic or icteric samples (serum or plasma) should only be tested with caution. Obviously contaminated samples should not be tested. Serum or plasma (EDTA, citrate, heparin) collected according to standard laboratory methods are suitable samples. Samples must not be thermally inactivated.

7.2.1 Dilution of Samples

Before running the test, all samples (V_1) must be diluted in dilution buffer (V_2) as follows:

West Nile Virus IgG ELISA:

$V_1 + V_2 = 1 + 100$	add	10 μ l	sample
	each to	1000 μ l	dilution buffer

After dilution and before pipetting into the microtiter plate the samples must be mixed thoroughly to prepare a homogenous solution.

West Nile Virus IgM ELISA:

Interference with rheumatoid factors

Rheumatoid factors are autoantibodies mainly of the IgM class, which preferably bind to IgG immune complexes. The presence of non-specific IgM antibodies (rheumatoid factors) can lead to false-positive results in the IgM assay. Furthermore, the possibility exists, that weak-binding pathogen-specific IgM antibodies may be displaced by stronger-binding IgG antibodies leading to a false negative IgM result. Therefore it is necessary to pretreat samples with rheumatoid factor-absorbents prior to IgM detection (Rf-Absorbent). Rf-absorption is performed by incubation of the sample in Rf-dilution buffer for 15 minutes at room temperature or over night at 4 °C. The test procedure is described in a separate instruction manual.

Before running the test, rheumatoid factor-absorbent (V_1) must be diluted 1+4 in dilution buffer (V_2).

$V_1 + V_2 = V_3 (1 + 4)$	add	200 μ l	Rf-absorbent
	each to	800 μ l	dilution buffer

Samples (V_4) must be diluted in this Rf-dilution buffer (V_3):

$V_4 + V_3 = 1 + 100$	add	10 μ l	sample
	each to	1000 μ l	Rf-dilution buffer

After dilution and before pipetting into the microtiter plate the samples must be mixed thoroughly to prepare a homogenous solution.

7.2.2 Sample Storage

The samples should not be stored for more than 7 days at 2 – 8 °C. Extended storage is possible at ≤ -20 °C. Avoid repeated freezing and thawing of samples. Diluted samples can be stored at 2 – 8 °C for one week.

7.3 Preparation of Kit Reagents

Bring all reagents to room temperature before testing.

7.3.1 Microtiter Test Strips

The microtiter test strips labeled with abbreviations for pathogen and immunoglobulin class are packed with a desiccant in an aluminum bag. To open the aluminum bag of the microtiter plate please cut off the top of the marked side only, in order to guarantee proper resealing. Take unrequired cavities out of the frame and put them back into the aluminum bag. Close bag carefully to ensure airtight conditions. Do not use the strips if the aluminum bag is damaged or if the bag with remaining strips and desiccant was not properly resealed.

7.3.2 Control Sera / Standard Sera

Control and standard sera are ready-to-use and must not be diluted any further. For each test run - independent of the number of microtiter test strips to be used - control and standard sera must be included. Standard and cut off sera should be set up in duplicate. Do not treat control sera with Rf-absorbent.

7.3.3 Anti-human IgG or IgM AP-Conjugate (ready-to-use)

Conjugates with the same concentration (+, ++, +++) and of the same immunoglobulin class are interchangeable. Please refer also to the specification on the label. Avoid contamination of ready-to-use conjugates e. g. by using sterile tips.

7.3.4 Washing Solution

Dilute washing buffer concentrate (V₁) 1:30 with aqua dest. to a final volume of V₂. Bottles used for the working dilution should be cleaned regularly. Discard cloudy solutions.

Example:

Buffer concentrate (V ₁)	Final volume (V ₂)
33.3 ml	1000 ml
1.0 ml	30 ml

7.3.5 Dilution Buffer for Samples (ready-to-use)

Discard cloudy solutions.

7.3.6 Substrate (ready-to-use)

Substrate in unopened bottle may have a slightly yellow coloring, which does not reduce the quality of the product! Avoid contamination.

7.3.7 Stopping Solution (ready-to-use)

7.4 Overview - Test Procedure

IBL-America West Nile Virus IgG/IgM

In case of IgM detection absorption of rheumatoid factor, see No. 7.2.1;
Incubation 15 minutes at room temperature or over night at 4°C

sample dilution¹
1+100

Pipette diluted samples and ready-to-use control /
standard sera into the microtest wells (100 µl)



INCUBATION 60 Min./ 37 °C
moist chamber



WASH (4 x 300 µl [DIL] [WASH])²



Pipette conjugate solution [APC] (100 µl)



INCUBATION 30 Min./ 37 °C
moist chamber



WASH (4 x 300 µl [DIL] [WASH])²



Pipette substrate solution [pNPP] (100 µl)



INCUBATION 30 Min./ 37 °C
moist chamber



Pipette stopping solution [STOP] (100 µl)



READ EXTINCTION at 405 nm

¹Special dilution buffers for the following IBL-America tests:
Borrelia burgdorferi IgG, IgM, EBV EA IgG

²For manual use:
tap plate at the end of the wash procedure on paper towel.

7.5 Manual Test Procedure

1. Place the required number of **cavities in the frame** and prepare a protocol sheet.
2. Add each **100 µl of diluted sample or ready-to-use controls** into the appropriate wells of microtiter test strips. Spare one well for substrate blank, e.g.:

well no.	IgG	IgM
A1	Substrate Blank	Substrate Blanks
B1	Negative Control	Negative Control
C1	Standard Serum	Cut-off Serum
D1	Standard Serum	Cut-off Serum
E1	Sample 1	Positive Control
F1	Sample 2	Sample 1

3. **Sample incubation** for 60 minutes (+/- 5 min) at 37 °C (+/- 1°C) in moist chamber
4. After incubation **wash** all wells with washing solution (by automated washer or manually):
 - aspirate or shake out the incubation solution
 - fill each well with 300 µl washing solution
 - aspirate or shake out the washing buffer
 - repeat the washing procedure 3 times (altogether 4 times!)
 - dry by tapping the microtiter plate on a paper towel
5. **Addition of conjugate**
Add 100 µl of the ready-to-use IgG/IgM conjugate to the appropriate wells (except substrate blank)
6. **Conjugate incubation** for 30 minutes (+/- 1 min)* at 37 °C (+/- 1 °C) in moist chamber.
7. After incubation **wash** all wells with washing solution (see above)
8. **Addition of substrate**
Add 100 µl of ready-to-use substrate solution to each well (including well for substrate blank!)
9. **Substrate incubation** for 30 minutes (+/- 1 min)* at 37 °C (+/- 1 °C) in moist chamber.
10. **Stopping of the reaction**
Add 100 µl stopping solution to each well, tap microtiter plate gently to mix.
11. **Read extinction**
Read optical density (OD) within 60 minutes at 405 nm against substrate blank, reference wave length between 620 nm and 690 nm (e.g. 650 nm).

* Please note, that under special working-conditions internal laboratory adaptations of the incubation times may be necessary.

7.6 Automated Test Procedure

The ELISA are validated for use with Immunomat (using the following consumables: VT124, VT111, VT112) and suited for processing on similar analyzers. For processing on the Immunomat the current software version including reagent check has to be used. The automated processing is performed analogous to manual use. Please note, that under special working-conditions internal laboratory adaptations of the substrate incubation times may be necessary.

7.7 Positive Control / Accuracy Control

For the periodic verification of the test method, in order to fulfil the requirements of laboratory internal quality management systems, we recommend using IBL-America ELISA controls (cat.-no. IB05xxxCON, see also chapter 5) to determine precision and accuracy of the test runs. The use of IBL-America ELISA controls is described in specific instruction manuals.

8 TEST EVALUATION

For qualitative interpretation of serum samples a lot specific correction factor as well as a lot specific grey zone is calculated by manufacturer for each kit lot. These values can be found on the lot specific quality certificate included in each test kit.

For test run control a standard serum is used in each individual test run. For this control serum a reference value with a validity range is determined by the quality control of the manufacturer. Within this range a correct cut-off interpretation is ensured.

8.1 Criteria of Validity

The substrate blank must be < 0.25 OD

The negative control must produce a negative test result.

The mean OD-value (after subtraction of the substrate blank!) of the standard serum must be within the validity range, which is given on the lot specific quality control certificate.

The variation of OD-values of the standard serum may not be higher than 20%.

If these criteria are not met, the test is not valid and must be repeated.

8.2 Cut-off Calculation

A lot specific quality control certificate is included in the test kit so that the obtained OD values can be interpreted qualitatively. The substrate blank must be subtracted from all OD values prior to evaluation.

To fix the cut-off ranges multiply the mean value of the measured standard OD with the lot specific correction factor from the quality control certificate. Then add and subtract the lot specific grey zone percentage mentioned on the quality certificate to obtain the upper and lower cut-off. The following numbers are an example only, the valid data you will find in the lot-specific QC certificate which comes with each kit.

Lot specific correction factor: 0.805

Lot specific grey zone: 15%

If the measured mean absorbance value of the standard serum is 0.84 OD, the range of the cut-off is:

Lower cut-off: $(0.84 * 0.805) - 15\% = \text{OD } 0.575$

Upper cut-off: $(0.84 * 0.805) + 15\% = \text{OD } 0.778$

8.3 Borderline Ranges

The borderline range indicates the range for borderline test results. Values obtained, when testing a sample, which fall below this range indicate a negative test result; values above the borderline range are interpreted positive. In cases where the results are within the borderline range a definitive interpretation of the result is not possible. In such cases, the test should be repeated in parallel with a follow-up sample taken one to two weeks later (serum pair).

9 PERFORMANCE CHARACTERISTICS

9.1 Sensitivity and Specificity

IBL-America West Nile Virus IgG and IgM

West Nile IgG Sensitivity is >92.6% and Specificity is >96.9%.

West Nile IgM Sensitivity is >99% and Specificity is > 99%.

9.2 Reproducibility

Intraassay reproducibility was determined by testing sera of different reactivities 20 times in one test run. Interassay reproducibility was determined by testing sera of different reactivities in 10 independent test runs on 5 different days.

$$\text{Coefficient of Variation (CV \%)} = \frac{\text{Standard deviation}}{\text{Mean value}} \times 100$$

West Nile Virus IgG:

Sample	Mean Value (OD)	Intraassay (CV %)	Mean Value (OD)	Interassay (CV %)
Sample 1	0.396	3.0	0.413	8.0
Sample 2	0.611	2.2	0.626	7.6
Sample 3	1.259	6.2	1.355	3.6

West Nile Virus IgM:

Sample	Mean Value (OD)	Intraassay (CV %)	Mean Value (OD)	Interassay (CV %)
Sample 1	0.532	3.7	0.434	14.2
Sample 2	1.035	5.0	0.888	6.0
Sample 3	1.542	2.1	1.394	3.6

9.3 Cross-reactivities

West Nile Virus IgG:

To determine detection of cross-reactive antibodies directed against different parameters sera were analyzed with the IBL-America West Nile Virus IgG ELISA and a commercially available anti-West Nile Virus IgG ELISA. Positive sera (10 sera each) for Dengue Virus IgG, Varicella Zoster Virus IgG, Cytomegalovirus IgG, Epstein-Barr Virus VCA IgG, Measles Virus IgG, Mumps Virus IgG, HSV 1 IgG, Rubella Virus IgG and Mycoplasma pneumoniae IgG have been tested as well as sera positive for rheumatoid factor (RF) and anti-nuclear antibodies (ANA). Within this internal evaluation potential cross-reactivities with one Dengue Virus IgG and one RF positive serum have been observed. Other crossreactivities cannot be ruled out in general.

West Nile Virus IgM:

To determine detection of cross-reactive antibodies directed against different parameters sera were analyzed with the IBL-America West Nile Virus IgM ELISA and a commercially available anti-West Nile Virus IgM ELISA. Positive sera (10 sera each) for Dengue Virus IgM, Varicella Zoster Virus IgM, Cytomegalovirus IgM, Epstein-Barr Virus VCA IgM, Measles Virus IgM, Mumps Virus IgM, HSV 1 IgM and Rubella Virus IgM have been tested as well as sera positive for rheumatoid factor (RF) and anti-nuclear antibodies (ANA). Within this internal evaluation potential cross-reactivities with one Measles Virus IgM positive serum sample has been observed. Other cross-reactivities cannot be ruled out in general.

9.4 Interfering Substances


West Nile Virus IgG/IgM:

To determine the influence of interfering substances, sera with different reactivities were analyzed with the West Nile Virus IgG/IgM ELISA. No interferences have been detected for sera with concentrations up to 2.00 g/L hemoglobin, 11.50 g/L lipemia/triglyceride or 0.201 g/L bilirubin (conjugated and unconjugated).

10 SAFETY MEASURES

10.1 Statements of Warning

The IBL-America test kits are designed for use by qualified personnel who are familiar with good laboratory practice. All kit reagents and specimens should be handled carefully, using established good laboratory practice.

- This kit contains human blood components. Although all control- and cut-off sera have been tested and found negative for anti-HIV-ab, HBs-Ag (*Hepatitis B-Virus-surface Antigen*) and anti-HCV-ab, they should be considered potentially infectious.
- Do not pipette by mouth.
- Do not smoke, eat or drink in areas in which specimens or kit reagents are handled.
- Wear disposable gloves, laboratory coat and safety glasses while handling kit reagents or specimens. Wash hands thoroughly afterwards.
- Sample material and other potentially infectious material should be decontaminated after the test run.
- Reagents should be stored safely and be inaccessible to unauthorized access e.g. children.
- Stopping solution:  corrosive (C); causes acid burn (R34)
Use safety glasses, gloves and laboratory coat while handling!

10.2 Disposal

Please observe the relevant statutory requirements!

11 REFERENCES

- [1] Beasley DWC., Holbrook MR., Travassos da Rosa A., Coffey L., Carrara AS., Philippi-Falkenstein K., Bohm Jr RP., Ratterree MS. Lillibridge KM., Ludwig GV., Estrada-Franco J., Weaver SC., Tesh RB., Shope RE., Barrett ADT. Use of a Recombinant Envelope Protein Subunit Antigen for Specific Serological Diagnosis of West Nile Virus Infection *Journal of Clinical Microbiology*; 42 2759-2765; 2004
- [2] Campell GL., Marfin AA., Lanciotti RS., Gubler DJ. West Nile Virus *Lancet Infect. Dis.*; 2 519-529; 2002
- [3] Chávez JH., Silva JR., Amarilla AA., Figueiredo LTM. Domain III peptides from flavivirus envelope protein are useful antigens for serologic diagnosis and targets for immunization *Biologicals*; 38 613-618; 2010
- [4] Dauphin G. & Zientara S. West Nile Virus: Recent trends in diagnosis and vaccine development *Vaccine*; 25 5563-5576; 2007
- [5] Diamond MS., Pierson TC., Fremont DH. The Structural Immunology of Antibody Protection against West Nile Virus *Immunol Rev.*; 225 212-225; 2008
- [6] Kemmerly SA. Diagnosis and Treatment of West Nile Infections *The Ochsner Journal*; 5 16-17; 2003
- [7] Monini M., Falcone E., Busani L., Romi R., Ruggeri FM. West Nile Virus: Characteristics of an African Virus Adapting to the Third Millennium World *The Open Virology Journal*; 4 42-51; 2010
- [8] Niedrig M., Sonnenberg K., Steinhagen K., Paweska JT. Comparison of ELISA and immunoassays for measurement of IgG and IgM antibody to West Nile Virus in human sera against virus neutralisation *Journal of Virological Methods*; 139 103-105; 2007
- [9] Oliphant T., Engle M., Nybakken GE., Doane C., Johnson S., Huang L., Gorlatov S., Mehlhop E., Marri A., Chung KM., Ebel GD., Kramer LD., Fremont DH., Diamond MS. Development of a Humanized monoclonal Antibody with Therapeutic Potential against West Nile Virus *Nat. Med.*; 11 522-530; 2005
- [10] Rossi SL., Ross TM., Evans JD. West Nile Virus *Clin Lab Med*; 30 47-65; 2010

Manufactured for Immuno-Biological Laboratories Inc.
(IBL-America)
8201 Central Avenue, NE, Suite P
Minneapolis, MN 55432
Tel: 763-780-2955
Toll Free: 1-888-523-1246

