

Product information

Information about other products is available at: www.ibl-america.com



User's Manual

Treponema pallidum (Syphilis) IgG ELISA


Enzyme immunoassay for the determination of IgG-class antibodies to *Treponema pallidum* in serum or plasma



IB79832



96 Wells



For Research Use Only – Not for Use in Diagnostic Procedures

Please use only the valid version of the Instructions for Use provided with the kit.

Table of Contents

1	INTRODUCTION	3
2	PRINCIPLE OF THE TEST	3
3	WARNINGS AND PRECAUTIONS	4
4	REAGENTS	5
5	SAMPLE	6
6	ASSAY PROCEDURE	7
7	RESULTS	8
8	QUALITY CONTROL	8
9	ASSAY CHARACTERISTICS	9
10	LIMITATIONS OF USE	10
11	LEGAL ASPECTS.....	10
12	REFERENCES / LITERATURE	10
	SHORT INSTRUCTIONS FOR USE	11
	SYMBOLS USED WITH IBL-AMERICA ASSAYS	12

1 INTRODUCTION

1.1 Intended Use

The **IBL America Treponema pallidum IgG Enzyme Immunoassay Kit** provides materials for the determination of IgG-class antibodies to *Treponema pallidum* in serum or plasma (EDTA-, heparin- or citrate plasma). **For Research Use Only – Not for Use in Diagnostic Procedures.**

1.2 Summary and Explanation

Spirochetes are motile bacteria with a periplasmic axial filament. All pathogenic species belong to the family Treponemataceae, which includes the three genera: *Treponema*, *Borrelia*, and *Leptospira*. The *Treponema* are motile bacteria, 5-15 μ in length and 0.2 μ in width, containing about 10 flexible, undulating, spiral shaped rods. *Treponema pallidum*, the causative agent of Syphilis, is transmitted by direct contact, usually through sexual intercourse. Syphilis along with Gonorrhoea, Chancroid and Lymphogranuloma venereum, designated as a venereal disease, or VD, is an acute and chronic infectious disease. After an incubation period of 12-30 days, the first symptoms to appear are chancres, soon followed by syphilitic ulcers which then spontaneously disappear in a few weeks. During this first stage (primary syphilis) the *Treponema pallidum* propagates in related lymph nodes to be distributed to the whole body stream. Three further stages of disease follow which are classified as secondary, tertiary, and quaternary syphilis.

2 PRINCIPLE OF THE TEST

The **IBL America Treponema pallidum IgG ELISA Kit** is a solid phase enzyme-linked immunosorbent assay (ELISA)

Microtiter wells as a solid phase are coated with *Treponema pallidum* antigen.

Diluted samples and **ready-for-use controls** are pipetted into these wells. During incubation *Treponema pallidum*-specific antibodies of positive samples and controls are bound to the immobilized antigens.

After a washing step to remove unbound sample and control material horseradish peroxidase conjugated anti-human IgG antibodies are dispensed into the wells. During a second incubation this anti-IgG conjugate binds specifically to IgG antibodies resulting in the formation of enzyme-linked immune complexes.

After a second washing step to remove unbound conjugate the immune complexes formed (in case of positive results) are detected by incubation with TMB substrate and development of a blue color. The blue color turns into yellow by stopping the enzymatic indicator reaction with sulfuric acid.

The intensity of this color is directly proportional to the amount of *Treponema pallidum*-specific IgG antibody in the sample. Absorbance at 450 nm is read using an ELISA microtiter plate reader.

3 WARNINGS AND PRECAUTIONS

- This kit is for research use only. Not for use in diagnostic procedures.
- Before starting the assay, read the instructions completely and carefully. Use the valid version of the package insert provided with the kit. Be sure that everything is understood.
- All reagents of this test kit which contain human serum or plasma have been tested and confirmed negative for HIV I/II, HBsAg and HCV by FDA approved procedures. All reagents, however, should be treated as potential biohazards in use and for disposal.
- Avoid contact with Stop Solution containing 0.2 mol/L H₂SO₄. It may cause skin irritation and burns.
- TMB substrate has an irritant effect on skin and mucosa. In case of possible contact, wash eyes with an abundant volume of water and skin with soap and abundant water. Wash contaminated objects before reusing them. If inhaled, take the person to open air.
- The microplate contains snap-off strips. Unused wells must be stored at 2 °C to 8 °C in the sealed foil pouch and used in the frame provided
- Pipetting of samples and reagents must be done as quickly as possible and in the same sequence for each step.
- Use reservoirs only for single reagents. This especially applies to the substrate reservoirs. Using a reservoir for dispensing a substrate solution that had previously been used for the conjugate solution may turn solution colored. Do not pour reagents back into vials as reagent contamination may occur.
- Mix the contents of the microplate wells thoroughly to ensure good test results. Do not reuse microwells.
- Do not let wells dry during assay; add reagents immediately after completing the rinsing steps.
- Allow the reagents to reach room temperature (21 °C to 26 °C) before starting the test. Temperature will affect the absorbance readings of the assay. However, values for the samples will not be affected.
- Never pipette by mouth and avoid contact of reagents and samples with skin and mucous membranes.
- Do not smoke, eat, drink or apply cosmetics in areas where samples or kit reagents are handled.
- Wear disposable latex gloves when handling samples and reagents. Microbial contamination of reagents or samples may give false results.
- Handling should be in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation.
- Do not use reagents beyond expiry date as shown on the kit labels.
- All indicated volumes have to be performed according to the protocol. Optimal test results are only obtained when using calibrated pipettes and microtiter plate readers.
- Do not mix or use components from kits with different lot numbers. It is advised not to exchange wells of different plates even of the same lot. The kits may have been shipped or stored under different conditions and the binding characteristics of the plates may result slightly different.
- Chemicals and prepared or used reagents have to be treated as hazardous waste according the national biohazard safety guideline or regulation.
- For information on hazardous substances included in the kit please refer to Safety Data Sheets. Safety Data Sheets for this product are available upon request directly from IBL America.

4 REAGENTS

4.1 Reagents provided

1. **SORB MT** **Microtiter wells**, 12 x 8 (break apart) strips, 96 wells; Wells coated with inactivated *Treponema pallidum* antigen. (incl. 1 strip holder and 1 cover foil)
2. **SAM DIL** **Sample Diluent** *, 1 vial, 100 mL, ready to use, colored yellow; pH 7.2 ± 0.2.
3. **CAL C** **Pos. Control** *, 1 vial, 2.0 mL, ready to use; colored yellow, red cap.
4. **CAL A** **Neg. Control** *, 1 vial, 2.0 mL, ready to use; colored yellow, yellow cap.
5. **CAL B** **Cut-off Control** *, 1 vial, 2.0 mL, ready to use; colored yellow, black cap.
6. **ENZ CONJ** **Enzyme Conjugate***, 1 vial, 20 mL, ready to use, colored red, antibody to human IgG conjugated to horseradish peroxidase.
7. **SUB TMB** **Substrate Solution**, 1 vial, 14 mL, ready to use, Tetramethylbenzidine (TMB).
8. **STOP SOLN** **Stop Solution**, 1 vial, 14 mL, ready to use, contains 0.2 mol/L H₂SO₄, Avoid contact with the stop solution. It may cause skin irritations and burns.
9. **WASH SOLN 20x** **Wash Solution** *, 1 vial, 30 mL (20X concentrated for 600 mL), pH 6.5 ± 0.1, see „Preparation of Reagents“.

* Contain non-mercury preservative.

4.1.1 Material required but not provided

- A microtiter plate calibrated reader (450/620 nm ±10 nm)
- Calibrated variable precision micropipettes
- Incubator 37 °C
- Manual or automatic equipment for rinsing wells
- Vortex tube mixer
- Deionised or (freshly) distilled water
- Timer
- Absorbent paper

4.2 Storage Conditions

When stored at 2 °C to 8 °C unopened reagents will retain reactivity until expiration date. Do not use reagents beyond this date.

Opened reagents must be stored at 2 °C to 8 °C. Microtiter wells must be stored at 2 °C to 8 °C. Once the foil bag has been opened, care should be taken to close it tightly again.

Opened kits retain activity for two months if stored as described above.

4.3 Reagent Preparation

Allow all reagents and required number of strips to reach room temperature prior to use.

Wash Solution

Dilute **Wash Solution 1+19** (e.g. 10 mL + 190 mL) with fresh and germ free redistilled water. This diluted wash solution has a pH value of 7.2 ± 0.2.

Consumption: ~ 5 mL per determination.

Crystals in the solution disappear by warming up to 37 °C in a water bath. Be sure that the crystals are completely dissolved before use.

The diluted Wash Solution is stable for 4 weeks at 2 °C to 8 °C.

4.4 Disposal of the Kit

The disposal of the kit must be made according to the national regulations. Special information for this product is given in the Material Safety Data Sheets.

4.5 Damaged Test Kits

In case of any severe damage to the test kit or components, IBL America has to be informed in writing, at the latest, one week after receiving the kit. Severely damaged single components should not be used for a test run. They have to be stored until a final solution has been found. After this, they should be disposed according to the official regulations.

5 SAMPLE

Serum or plasma (EDTA-, heparin- or citrate plasma) can be used in this assay. Do not use haemolytic, icteric or lipaemic samples.

5.1 Sample Collection

Serum:

Collect blood by venipuncture (e.g. Sarstedt Monovette for serum), allow to clot, and separate serum by centrifugation at room temperature. Do not centrifuge before complete clotting has occurred. Subjects receiving anticoagulants may require increased clotting time.

Plasma:

Whole blood should be collected into centrifuge tubes containing anti-coagulant (e.g. Sarstedt Monovette with the appropriate plasma preparation) and centrifuged immediately after collection.

5.2 Sample Storage

Specimens should be capped and may be stored for up to 24 hours at 2 °C to 8 °C prior to assaying. Specimens held for a longer time should be frozen only once at –20 °C prior to assay. Thawed samples should be inverted several times prior to testing.

5.3 Sample Dilution

Prior to assaying dilute each sample **1+100** with *Sample Diluent*, e.g. 10 µL of sample + 1 mL of *Sample Diluent* (mix well).

Let stand for at least 15 minutes and mix well again.

Please note: Controls are ready for use and must not be diluted!

6 ASSAY PROCEDURE

6.1 General Remarks

- **It is very important to bring all reagents, samples and controls to room temperature before starting the test run!**
- Once the test has been started, all steps should be completed without interruption.
- Use new disposal plastic pipette tips for each standard, control or sample in order to avoid cross contamination
- Absorbance is a function of the incubation time and temperature. Before starting the assay, it is recommended that all reagents are ready, caps removed, all needed wells secured in holder, etc. This will ensure equal elapsed time for each pipetting step without interruption.
- As a general rule the enzymatic reaction is linearly proportional to time and temperature.
- Close reagent vials tightly immediately after use to avoid evaporation and microbial contamination.
- To avoid cross-contamination and falsely elevated results pipette samples and dispense conjugate without splashing accurately to the bottom of wells.
- During incubation cover microtiter strips with foil to avoid evaporation.

6.2 Test Procedure

Prior to commencing the assay, the **distribution and identification plan** for all samples and controls should be carefully established on a form supplied in the kit.

1. Select the required number of microtiter strips or wells and insert them into the holder.
Please allocate at least:

1 well	(e.g. A1)	for the substrate blank,	
1 well	(e.g. B1)	for the <i>Neg. Control</i> ,	
2 wells	(e.g. C1+D1)	for the <i>Cut-off Control</i>	and
1 well	(e.g. E1)	for the <i>Pos. Control</i> .	

 It is left to the user to determine controls and samples in duplicate.
2. Dispense

100 µL of <i>Neg. Control</i>	into well	B1
100 µL of <i>Cut-off Control</i>	into wells	C1 and D1
100 µL of <i>Pos. Control</i>	into well	E1

 and
100 µL of each diluted sample with new disposable tips into appropriate wells.
 Leave well A1 for substrate blank!
3. Cover wells with foil supplied in the kit. Incubate for **60 minutes at 37 °C**.
4. Briskly shake out the contents of the wells. Rinse the wells 5 times with diluted *Wash Solution* (**300 µL per well**). Strike the wells sharply on absorbent paper to remove residual droplets.
Important note: The sensitivity and precision of this assay is markedly influenced by the correct performance of the washing procedure!
5. Dispense **100 µL Enzyme Conjugate** into each well, **except A1**.
6. Incubate for **30 minutes at room temperature (20 °C to 25 °C)**. *Do not expose to direct sun light!*
7. Briskly shake out the contents of the wells. Rinse the wells 5 times with diluted *Wash Solution* (300 µL per well). Strike the wells sharply on absorbent paper to remove residual droplets.
8. Add **100 µL of Substrate Solution** into all wells.
9. Incubate for **exactly 15 minutes at room temperature (20 °C to 25 °C) in the dark**.
10. Stop the enzymatic reaction by adding **100 µL of Stop Solution** to each well. Any blue color developed during the incubation turns into yellow.
Note: Highly positive samples can cause dark precipitates of the chromogen!
11. Read the optical density at **450/620 nm** with a microtiter plate reader **within 30 minutes** after adding the *Stop Solution*.

6.3 Measurement

Adjust the ELISA microplate or microstrip reader **to zero** using the **substrate blank in well A1**.
 If - due to technical reasons - the ELISA reader cannot be adjusted to zero using the substrate blank in well A1, subtract the absorbance value of well A1 from all other absorbance values measured in order to obtain reliable results! **Measure the absorbance** of all wells **at 450 nm** and record the absorbance values for each control and sample in the distribution and identification plan.
Dual wavelength reading using 620 nm as reference wavelength is recommended.
 Where applicable **calculate the mean absorbance values** of all duplicates.

7 RESULTS

7.1 Validation of the Test Run

The test run may be considered valid provided the following criteria are met:

Substrate blank in A1:	Absorbance value lower than 0.100
Neg. Control in B1:	Absorbance value lower than 0.200
Cut-off Control in C1/D1 :	Absorbance value between 0.350 – 0.850
Pos. Control in E1	Absorbance value between 0.650 – 3.000

7.2 Calculation

Mean absorbance value of Cut-off Control [CO]

Calculate the mean absorbance value of the two (2) Cut-off Control determinations (e.g. in C1/D1).

Example: $(0.49 + 0.51) / 2 = 0.50 = CO$

7.3 Results

POSITIVE	Sample (mean) absorbance values more than 10 % above CO (Mean OD _{sample} > 1.1 x CO)
GREY ZONE	Sample (mean) absorbance values from 10 % above to 10 % below CO repeat test 2 - 4 weeks later - with <u>new</u> samples ($0.9 \times CO \leq \text{Mean OD}_{\text{sample}} \leq 1.1 \times CO$)
NEGATIVE	Results in the second test again in the grey zone \Rightarrow NEGATIVE Sample (mean) absorbance values more than 10 % below CO (Mean OD _{sample} < 0.9 x CO)

7.3.1 Results in Units [U]

$$\frac{\text{Sample (mean) absorbance value} \times 10}{CO} = [\text{Units} = U]$$

Example:
$$\frac{1.580 \times 10}{0.50} = 32 U$$

Results

Cut-off value:	10	DU
Grey zone:	9 - 11	DU
Negative:	< 9	DU
Positive:	> 11	DU

8 QUALITY CONTROL

It is recommended to use control samples according to state and federal regulations. The use of control samples is advised to assure the day to day validity of results. Use controls at both normal and pathological levels. It is also recommended to make use of national or international Quality Assessment programs in order to ensure the accuracy of the results. If the results of the assay do not fit to the established acceptable ranges of control materials results should be considered invalid. In this case, please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods. After checking the above mentioned items without finding any error contact your distributor or IBL America directly.

9 ASSAY CHARACTERISTICS

9.1 Assay Dynamic Range

The range of the assay is between 0.69 U/mL - 60 U/mL.

9.2 Analytical Sensitivity

The sensitivity of the IBL-AMERICA ELISA was calculated by adding 2 standard deviations from the mean of 20 replicate analyses of the negative control and was found to be 0.69 U/mL (OD₄₅₀ = 0.042).

9.3 Specificity of Antigen

The antigen used for the IBL-AMERICA Treponema pallidum IgG ELISA shows no cross-reactivity to Fasciola, Strongyloides, Ascaris, Schistosoma, Trichinella and Giardia lamblia. None of the following samples with interference factors will interfere with the ELISA: samples with HAMA or different ANA, samples with tumor marker (CYFRA), and samples with pregnancy hormones.

9.4 Specificity

The specificity is defined as the probability of the assay of scoring negative in the absence of the specific analyte. It is 100%.

9.5 Sensitivity

The sensitivity is defined as the probability of the assay of scoring positive in the presence of the specific analyte. It is 100%.

9.6 Reproducibility

9.6.1 The intra-assay

The intra-assay (within-run) precision of the IBL-AMERICA ELISA was determined by 20 x measurements of 12 samples covering the measuring range of the ELISA

Sample	Mean OD	Intra-Assay CV (%)	n
1	0.14	4.64	20
2	0.43	2.75	20
3	0.51	4.74	20
4	0.99	7.09	20
5	0.94	2.63	20
6	0.68	5.14	20
7	1.13	4.32	20
8	1.05	3.81	20
9	1.48	6.99	20
10	2.15	6.67	20
11	1.60	7.37	20
12	2.57	6.81	20

9.6.2 The inter-assay

The inter-assay variation of the IBL-AMERICA ELISA was determined with cut-off control, positive control, and 3 samples with 2 production kits in 10 independent runs with 2 replicates per run.

Sample	Mean OD	Inter-Assay CV (%)	n
1	2.15	3.07	40
2	2.51	2.25	40
3	0.97	3.21	40

10 LIMITATIONS OF USE

Bacterial contamination or repeated freeze-thaw cycles of the sample may affect the absorbance values. In immune-compromised samples and newborns serological data only have restricted value.

10.1 Interfering Substances

Haemoglobin (up to 4 mg/mL), Bilirubin (up to 0.5 mg/mL) and Triglyceride (up to 30 mg/mL) have no influence on the assay results.

11 LEGAL ASPECTS

11.1 Reliability of Results

The test must be performed exactly as per the manufacturer's instructions for use. Moreover the user must strictly adhere to the rules of GLP (Good Laboratory Practice) or other applicable national standards and/or laws. This is especially relevant for the use of control reagents. It is important to always include, within the test procedure, a sufficient number of controls for validating the accuracy and precision of the test.

The test results are valid only if all controls are within the specified ranges and if all other test parameters are also within the given assay specifications. In case of any doubt or concern please contact IBL America.

11.2 Liability

















Any modification of the test kit and/or exchange or mixture of any components of different lots from one test kit to another could negatively affect the intended results and validity of the overall test. Such modification and/or exchanges invalidate any claim for replacement.

Claims submitted due to customer misinterpretation of laboratory results are also invalid. Regardless, in the event of any claim, the manufacturer's liability is not to exceed the value of the test kit. Any damage caused to the test kit during transportation is not subject to the liability of the manufacturer.











12 REFERENCES / LITERATURE

1. Luger, A., B.L. Schmidt, und F. Gschnait. 1983. Neue Fortschritte der Syphilisserologie. *Wr. Klein. Wsch.* 95: 440-443
2. Penn. C. W., M. J. Baily, and Cockayne. 1985. The axial filament antigen of *Treponema pallidum*. *Immunology* 54: 635-641

SHORT INSTRUCTIONS FOR USE

	All reagents and samples must be allowed to come to room temperature (18-25°C) before use.
	Leave well A1 for substrate Blank. Dispense 100 µl of Controls into appropriate wells.
	Dispense 100 µl of sample into selected wells.
	Cover wells with foil. Incubate for 60 minutes at 37 °C.
	Briskly shake out the contents of the wells.
	Rinse the wells 5 times with diluted Wash Solution (300 µl per well).
	Strike the wells sharply on absorbent paper to remove residual droplets.
	Dispense 100 µl of Enzyme-Conjugate into each well.
	Incubate for 30 minutes at room temperature.
	Briskly shake out the contents of the wells.
	Rinse the wells 5 times with diluted Wash Solution (300 µl per well).
	Strike the wells sharply on absorbent paper to remove residual droplets.
	Add 100 µl of Substrate Solution to each well.
	Incubate for 15 minutes at room temperature.
	Stop the reaction by adding 100 µl of Stop Solution to each well.
	Determine the absorbance of each well at 450 nm.

SYMBOLS USED WITH IBL-AMERICA ASSAYS

Symbol	English	Deutsch	Français	Espanol	Italiano
	European Conformity	CE-Konformitätskennzeichnung	Conforme aux normes européennes	Conformidad europea	Conformità europea
	Consult instructions for use	Gebrauchsanweisung beachten	Consulter les instructions d'utilisation	Consulte las Instrucciones	Consultare le istruzioni per l'uso
	In vitro diagnostic device	In-vitro-Diagnostikum	Usage Diagnostic in vitro	Diagnóstico in vitro	Per uso Diagnostica in vitro
	For research use only	Nur für Forschungszwecke	Seulement dans le cadre de recherches	Sólo para uso en investigación	Solo a scopo di ricerca
	Catalogue number	Katalog-Nr.	Référence	Número de catálogo	No. di Cat.
	Lot. No. / Batch code	Chargen-Nr.	No. de lot	Número de lote	Lotto no
	Contains sufficient for <n> tests/	Ausreichend für "n" Ansätze	Contenu suffisant pour "n" tests	Contenido suficiente para <n> ensayos	Contenuto sufficiente per "n" saggi
	Note warnings and precautions	Warnhinweise und Vorsichtsmaßnahmen beachten	Avertissements et mesures de précaution font attention	Tiene en cuenta advertencias y precauciones	Annoti avvisi e le precauzioni
	Storage Temperature	Lagerungstemperatur	Temperature de conservation	Temperatura de conservación	Temperatura di conservazione
	Expiration Date	Mindesthaltbarkeitsdatum	Date limite d'utilisation	Fecha de caducidad	Data di scadenza
	Legal Manufacturer	Hersteller	Fabricant	Fabricante	Fabbricante
<i>Distributed by</i>	Distributor	Vertreiber	Distributeur	Distribuidor	Distributore