

25OH Vitamin D Total ELISA

KAR1971

For Research Use Only. Not For Use in Diagnostic Procedures

RUO

For Research Use Only – Not for Use in Diagnostic Procedures





Distributed By:
IBL-America
8201 Central Ave NE, Suite P
Minneapolis, MN 55432
(888) 523-1246
www.ibl-america.com
info@ibl-america.com

For Research Use Only. Not For Use in Diagnostic Procedures

History

Summary of change:

Previous Version: 200224-1	Current Version: 220516																																																																																																								
Old DIAsource logo 	New DIAsource logo 																																																																																																								
Immunoenzymetric assay for the <i>in vitro</i> quantitative measurement of 25-hydroxyvitamin D2 and D3 (25OH-D2 and 25OH-D3) in serum.	Immunoenzymetric assay for the <i>in vitro</i> quantitative measurement of 25-hydroxyvitamin D2 and D3 (25OH-D2 and 25OH-D3) in serum and plasma .																																																																																																								
V. REAGENTS PROVIDED	V. REAGENTS PROVIDED Removal of the column relative to color code.																																																																																																								
<table border="1"> <tr> <td>INC</td> <td>BUF</td> <td>1 vial</td> <td>green</td> <td>Ready for use</td> </tr> <tr> <td colspan="2">Incubation Buffer with casein and proclin</td> <td>20 ml</td> <td></td> <td></td> </tr> <tr> <td>CHROM</td> <td>TMB</td> <td>1 vial</td> <td>orange</td> <td>Ready for use</td> </tr> <tr> <td colspan="2">Chromogenic solution TMB (Tetramethylbenzidine)</td> <td>12 ml</td> <td></td> <td></td> </tr> <tr> <td>STOP</td> <td>SOLN</td> <td>1 vial</td> <td></td> <td>Ready for use</td> </tr> <tr> <td colspan="2">Stop solution HCl 1M</td> <td>12 ml</td> <td></td> <td></td> </tr> </table>	INC	BUF	1 vial	green	Ready for use	Incubation Buffer with casein and proclin		20 ml			CHROM	TMB	1 vial	orange	Ready for use	Chromogenic solution TMB (Tetramethylbenzidine)		12 ml			STOP	SOLN	1 vial		Ready for use	Stop solution HCl 1M		12 ml			<table border="1"> <tr> <td>INC</td> <td>BUF</td> <td>1 vial</td> <td>Ready for use</td> </tr> <tr> <td colspan="2">Incubation Buffer with casein and proclin</td> <td>30 ml</td> <td></td> </tr> <tr> <td>CHROM</td> <td>TMB</td> <td>1 vial</td> <td>Ready for use</td> </tr> <tr> <td colspan="2">Chromogenic solution TMB (Tetramethylbenzidine)</td> <td>13 ml</td> <td></td> </tr> <tr> <td>STOP</td> <td>SOLN</td> <td>1 vial</td> <td>Ready for use</td> </tr> <tr> <td colspan="2">Stop solution 0.2M H₂SO₄</td> <td>13 ml</td> <td></td> </tr> </table>	INC	BUF	1 vial	Ready for use	Incubation Buffer with casein and proclin		30 ml		CHROM	TMB	1 vial	Ready for use	Chromogenic solution TMB (Tetramethylbenzidine)		13 ml		STOP	SOLN	1 vial	Ready for use	Stop solution 0.2M H ₂ SO ₄		13 ml																																																			
INC	BUF	1 vial	green	Ready for use																																																																																																					
Incubation Buffer with casein and proclin		20 ml																																																																																																							
CHROM	TMB	1 vial	orange	Ready for use																																																																																																					
Chromogenic solution TMB (Tetramethylbenzidine)		12 ml																																																																																																							
STOP	SOLN	1 vial		Ready for use																																																																																																					
Stop solution HCl 1M		12 ml																																																																																																							
INC	BUF	1 vial	Ready for use																																																																																																						
Incubation Buffer with casein and proclin		30 ml																																																																																																							
CHROM	TMB	1 vial	Ready for use																																																																																																						
Chromogenic solution TMB (Tetramethylbenzidine)		13 ml																																																																																																							
STOP	SOLN	1 vial	Ready for use																																																																																																						
Stop solution 0.2M H ₂ SO ₄		13 ml																																																																																																							
VII. REAGENT PREPARATION	VII. REAGENT PREPARATION																																																																																																								
D. Working HRP conjugate solution Prepare the solution according to the number of used strips, as indicated in the below table: for example for 6 strips (48 wells): 100 µl of concentrated conjugate and 50 µl of concentrated HRP to 10 ml of conjugate buffer.	D. Working HRP conjugate solution Prepare the solution according to the number of used strips, as indicated in the below table: for example for 6 strips (48 wells): 130 µl of concentrated conjugate and 65 µl of concentrated HRP to 13 ml of conjugate buffer.																																																																																																								
<table border="1"> <thead> <tr> <th>Nb. of strips</th> <th>Volume of Conjugate Buffer (ml)</th> <th>Volume of Concentrated Conjugate (µl)</th> <th>Volume of Concentrated HRP(µl)</th> </tr> </thead> <tbody> <tr><td>1</td><td>3</td><td>30</td><td>15</td></tr> <tr><td>2</td><td>5</td><td>50</td><td>25</td></tr> <tr><td>3</td><td>6</td><td>60</td><td>30</td></tr> <tr><td>4</td><td>8</td><td>80</td><td>40</td></tr> <tr><td>5</td><td>9</td><td>90</td><td>45</td></tr> <tr><td>6</td><td>10</td><td>100</td><td>50</td></tr> <tr><td>7</td><td>12</td><td>120</td><td>60</td></tr> <tr><td>8</td><td>14</td><td>140</td><td>70</td></tr> <tr><td>9</td><td>16</td><td>160</td><td>80</td></tr> <tr><td>10</td><td>18</td><td>180</td><td>90</td></tr> <tr><td>11</td><td>20</td><td>200</td><td>100</td></tr> <tr><td>12</td><td>22</td><td>220</td><td>110</td></tr> </tbody> </table>	Nb. of strips	Volume of Conjugate Buffer (ml)	Volume of Concentrated Conjugate (µl)	Volume of Concentrated HRP(µl)	1	3	30	15	2	5	50	25	3	6	60	30	4	8	80	40	5	9	90	45	6	10	100	50	7	12	120	60	8	14	140	70	9	16	160	80	10	18	180	90	11	20	200	100	12	22	220	110	<table border="1"> <thead> <tr> <th>Nb of strips</th> <th>Volume of Conjugate Buffer (ml)</th> <th>Volume of Concentrated Conjugate (µl)</th> <th>Volume of Concentrated HRP (µl)</th> </tr> </thead> <tbody> <tr><td>1</td><td>3</td><td>30</td><td>15</td></tr> <tr><td>2</td><td>5</td><td>50</td><td>25</td></tr> <tr><td>3</td><td>7</td><td>70</td><td>35</td></tr> <tr><td>4</td><td>9</td><td>90</td><td>45</td></tr> <tr><td>5</td><td>11</td><td>110</td><td>55</td></tr> <tr><td>6</td><td>13</td><td>130</td><td>65</td></tr> <tr><td>7</td><td>15</td><td>150</td><td>75</td></tr> <tr><td>8</td><td>17</td><td>170</td><td>85</td></tr> <tr><td>9</td><td>19</td><td>190</td><td>95</td></tr> <tr><td>10</td><td>21</td><td>210</td><td>105</td></tr> <tr><td>11</td><td>23</td><td>230</td><td>115</td></tr> <tr><td>12</td><td>25</td><td>250</td><td>125</td></tr> </tbody> </table>	Nb of strips	Volume of Conjugate Buffer (ml)	Volume of Concentrated Conjugate (µl)	Volume of Concentrated HRP (µl)	1	3	30	15	2	5	50	25	3	7	70	35	4	9	90	45	5	11	110	55	6	13	130	65	7	15	150	75	8	17	170	85	9	19	190	95	10	21	210	105	11	23	230	115	12	25	250	125
Nb. of strips	Volume of Conjugate Buffer (ml)	Volume of Concentrated Conjugate (µl)	Volume of Concentrated HRP(µl)																																																																																																						
1	3	30	15																																																																																																						
2	5	50	25																																																																																																						
3	6	60	30																																																																																																						
4	8	80	40																																																																																																						
5	9	90	45																																																																																																						
6	10	100	50																																																																																																						
7	12	120	60																																																																																																						
8	14	140	70																																																																																																						
9	16	160	80																																																																																																						
10	18	180	90																																																																																																						
11	20	200	100																																																																																																						
12	22	220	110																																																																																																						
Nb of strips	Volume of Conjugate Buffer (ml)	Volume of Concentrated Conjugate (µl)	Volume of Concentrated HRP (µl)																																																																																																						
1	3	30	15																																																																																																						
2	5	50	25																																																																																																						
3	7	70	35																																																																																																						
4	9	90	45																																																																																																						
5	11	110	55																																																																																																						
6	13	130	65																																																																																																						
7	15	150	75																																																																																																						
8	17	170	85																																																																																																						
9	19	190	95																																																																																																						
10	21	210	105																																																																																																						
11	23	230	115																																																																																																						
12	25	250	125																																																																																																						
IX. SAMPLE COLLECTION AND PREPARATION	IX. SAMPLE COLLECTION AND PREPARATION																																																																																																								
<ul style="list-style-type: none"> - This kit is suitable for serum samples. - Serum samples must be kept at 2-8°C. - If the test is not run within 24 hrs, sampling and storage at -20°C is recommended. 	<ul style="list-style-type: none"> - This kit is suitable for serum and heparinized plasma samples. - Serum and heparinized plasma samples must be kept at 2-8°C. - If the test is not run within 24 hrs, sampling and storage at -20°C is recommended. 																																																																																																								

- Avoid subsequent freeze-thaw cycles.

- Avoid subsequent freeze-thaw cycles.
 - Serum and heparinized plasma provide similar results. $Y(\text{Heparin plasma}) = 0.9922 \times (\text{serum}) + 0.2129 \text{ ng/ml}$, $R^2 = 0.9944$, $n = 10$

X. PROCEDURE

B. Procedure

3. Pipette 50 µl of each Calibrator, Control and Sample into the appropriate wells.
4. Pipette 150 µl of Incubation Buffer into all the wells.
8. Pipette 200 µl of the working HRP conjugate solution into each well. Incubate the microtiterplate for 30 minutes at room temperature, on a plate shaker (400 rpm)

X. PROCEDURE

B. Procedure

3. Pipette 25 µl of each Calibrator, Control and Sample into the appropriate wells.
4. Pipette 250 µl of Incubation Buffer into all the wells.
8. Pipette 250 µl of the working HRP conjugate solution into each well. Incubate the microtiterplate for 30 minutes at room temperature, on a plate shaker (400 rpm)

XII. TYPICAL DATA

25OH-ELISA		OD units
Calibrator	0 ng/ml	2.66
	5.3 ng/ml	2.39
	15.0 ng/ml	1.83
	25.7 ng/ml	1.46
	54.3 ng/ml	0.81
	133 ng/ml	0.21

XII. TYPICAL DATA

25OH-ELISA		OD units
Calibrator	0 ng/ml	2.79
	3.44 ng/ml	2.56
	16.28 ng/ml	1.93
	32.39 ng/ml	1.20
	63.54 ng/ml	0.49
	122.76 ng/ml	0.16

XIII. PERFORMANCE AND LIMITATIONS

A. Limits of Detection

The Limit of Blank (LoB), Limit of Detection (LoD), and the Limit of Quantitation (LoQ), were determined in accordance with the CLSI guideline EP17-A. The LoB was calculated by measuring the blank several times and calculating the 95th percentile of the distribution of the test values. The LoB was calculated to be 1.69ng/ml. The LoD was calculated as described in the guideline. The LoD was calculated to be 2.81ng/ml. The LoQ was calculated by testing 5 samples of low value 14 times in different test. The LoQ was calculated to be 4.39ng/ml with CV of 20%.

B. Specificity

Compound and Concentration	% Cross reaction
25OH-Vitamin D ₃ at 10 ng/mL	100
25OH-Vitamin D ₂ at 10 ng/mL	86
1,25(OH) ₂ -Vitamin D ₃ at 200 ng/mL	20
1,25(OH) ₂ -Vitamin D ₂ at 690 ng/mL	1.9
Vitamin D ₃ at 200 ng/mL	2.9
Vitamin D ₂ at 200 ng/mL	1.3
24,25(OH) ₂ -Vitamin D ₃ at 20 ng/mL	>100
25,26(OH) ₂ -Vitamin D ₃ at 4 ng/mL	>100
3-epi-25OH-Vitamin D ₃ at 20 µg/mL	0.1

The effect of potential interfering substances on samples using the DAsoure 25 OH Vitamin D Total ELISA test was evaluated. Different levels of Hemoglobin, Triglyceride, Vitamin C, Bilirubin Conjugate and Unconjugated and Zemplar in serum samples were tested on samples with different 25OH Vitamin D Concentration. Our acceptance criteria was to have interference of less than 10%. The tested substances did not affect the performance of the DAsoure 25 OH Vitamin D Total ELISA test.

XIII. PERFORMANCE AND LIMITATIONS

A. Limits of Detection

The Limit of Blank (LoB), Limit of Detection (LoD), and the Limit of Quantitation (LoQ), were determined in accordance with the CLSI guideline EP17-A. The LoB was calculated by measuring the blank several times and calculating the 95th percentile of the distribution of the test values. The LoB was calculated to be 2.07ng/ml. The LoD was calculated as described in the guideline. The LoD was calculated to be 3.26ng/ml. The LoQ was calculated by testing 5 samples of low value 10 times in different test. The LoQ was calculated to be 3.35ng/ml with CV of 20%.

B. Specificity

Compound and Concentration	% Cross reaction
25OH-Vitamin D ₃ at 10 ng/mL	92
25OH-Vitamin D ₂ at 10 ng/mL	91
1,25(OH) ₂ -Vitamin D ₃ at 200 ng/mL	3.10
1,25(OH) ₂ -Vitamin D ₂ at 667 ng/mL	0.35
Vitamin D ₃ at 200 ng/mL	0.17
Vitamin D ₂ at 200 ng/mL	0.22
3-epi-25OH-Vitamin D ₃ at 20 µg/mL	0.91

The effect of potential interfering substances on samples using the DAsoure 25 OH Vitamin D Total ELISA test was evaluated. Different levels of Hemoglobin, Triglyceride, Vitamin C, Bilirubin Conjugate and Unconjugated in serum samples were tested on samples with different 25OH Vitamin D Concentration. Our acceptance criteria was to have interference of less than 10%. The tested substances did not affect the performance of the DAsoure 25 OH Vitamin D Total ELISA test.

Substance	25OH Vitamin D (ng/mL)	Concentration of Interferent (mg/dL)	Mean % Variation	
Hemoglobin	7.6	250	-0.6%	
		500		
	29.3	250		
500				
Bilirubin Conjugated	6.0	50		-3.4%
		100		
	21.6	50		
100				
Bilirubin Unconjugated	7.6	50	2.5%	
		100		
	29.3	50		
100				
Triglyceride	7.6	7.5		-4.3%
		125		
	29.3	250		
500				
Vitamin C	6.0	7.5	-4.3%	
		125		
	29.3	250		
500				
Biotin	8.7	0.2		4.7%
		2		
	19.8	4		
0.2				
Zemplar	17.6	2	-4.4%	
		4		
	36.1	0.2		
2				
Hemoglobin	7.6	0.0013		-4.4%
		0.0025		
	29.3	0.0050		
0.0013				
Bilirubin Conjugated	6.0	0.0025	-4.4%	
		0.0050		
	21.6	0.0013		
0.0025				
Bilirubin Unconjugated	7.6	0.0013		-4.4%
		0.0025		
	29.3	0.0050		
0.0013				

Substance	25OH Vitamin D (ng/mL)	Concentration of Interferent (mg/dL)	Mean % Variation	
Hemoglobin	21.87	250	-3.7%	
		500		
	37.36	250		
500				
Bilirubin Conjugated	21.87	50		2.5%
		100		
	37.36	50		
100				
Bilirubin Unconjugated	21.87	50	-0.2%	
		100		
	37.36	50		
100				
Triglyceride	21.87	6.25		-2.1%
		125		
	37.36	250		
500				
Vitamin C	21.87	6.25	2.9%	
		125		
	37.36	250		
500				
Biotin	21.87	1		2.5%
		10		
	37.36	100		
100				
Hemoglobin	21.87	0.2	2.5%	
		2		
	37.66	4		
0.2				
Bilirubin Conjugated	6.0	2		2.5%
		4		
	21.6	0.2		
2				
Bilirubin Unconjugated	7.6	0.2	2.5%	
		4		
	29.3	0.2		
4				

C. Precision							
INTRA-ASSAY				INTER-ASSAY			
Sample	N	<X> ± SD (ng/ml)	C.V. (%)	Sample	N	<X> ± SD (ng/ml)	C.V. (%)
A	24	5.5 ± 0.4	7.8	A	39	17.7 ± 1.3	7.4
B	35	27.4 ± 1.6	5.7	B	10	26.3 ± 1.2	4.7
C	35	43.0 ± 1.2	2.7	C	10	42.1 ± 1.8	4.3
D	24	81.2 ± 2.0	2.5	D	21	85.4 ± 7.8	9.2

SD: Standard Deviation, CV: Coefficient of variation

C. Precision							
INTRA-ASSAY				INTER-ASSAY			
Sample	N	<X> ± SD (ng/ml)	C.V. (%)	Sample	N	<X> ± SD (ng/ml)	C.V. (%)
A	24	20.40 ± 0.69	3.4	A	10	12.62 ± 0.79	6.3
B	24	33.32 ± 0.96	2.9	B	10	20.99 ± 0.69	3.3
				C	10	33.22 ± 1.30	3.9
				D	10	70.25 ± 2.42	3.4

SD: Standard Deviation, CV: Coefficient of variation

E. Accuracy

RECOVERY TEST	
Added 25OH-Vit.D ₃ (ng/ml)	Recovery (%)
0	100
25	96
50	92
Added 25OH-Vit.D ₂ (ng/ml)	Recovery (%)
0	100
25	105
50	95

E. Accuracy

RECOVERY TEST	
Added 25OH-Vit.D ₃ (ng/ml)	Recovery (%)
5	90
10	92
25	85
50	71
Added 25OH-Vit.D ₂ (ng/ml)	Recovery (%)
5	88
10	91
25	88
50	83

DILUTION TEST

Sample Dilution	Theoretical Concentration (ng/mL)	Measured Concentration (ng/mL)	Slope	Y-Intercept	R ²	Recovery (%)
1/1	101.8	101.8	1.02	-1.91	>0.98	100
1/2	64.4	62.9				98
1/4	45.7	52.0				114
1/8	36.4	34.8				96
1/16	31.7	33.6				106

DILUTION TEST

DILUTION TEST:			
Sample dilution	Theoretical concentration (ng/ml)	Measured concentration (ng/ml)	Recovery (%)
1/1	-	65.20	-
1/2	32.60	29.34	90.0
1/4	16.30	15.01	92.1
1/8	8.15	8.91	109.3

Samples were diluted with the zero calibrator.

XVIII. SUMMARY OF THE PROTOCOL

	CALIBRATORS	SAMPLE(S)
	(μl)	CONTROLS (μl)
Calibrators (0-5)	50	-
Controls, Samples	-	50
Incubation Buffer	150	150
Working HRP Conjugate	200	200

XVIII. SUMMARY OF THE PROTOCOL

	CALIBRATORS	SAMPLE(S)
	(μl)	CONTROLS (μl)
Calibrators (0-5)	25	-
Controls, Samples	-	25
Incubation Buffer	250	250
Working HRP Conjugate	250	250

Read entire protocol before use.

25OH Vitamin D Total ELISA

I. INTENDED USE

Immunoenzymetric assay for the quantitative measurement of 25-hydroxyvitamin D₂ and D₃ (25OH-D₂ and 25OH-D₃) in serum and plasma.

II. GENERAL INFORMATION

- A. Proprietary name:** DIAsource 25OH Vitamin D Total ELISA Kit
- B. Catalog number:** KAR1971: 96 tests
- C. Manufactured by:** DIAsource ImmunoAssays S.A.
Rue du Bosquet, 2 B-1348 Louvain-la Neuve, Belgium.

III. BACKGROUND

Vitamin D is the generic term used to designate Vitamin D₂ or ergocalciferol and Vitamin D₃ or cholecalciferol.

Humans naturally produce Vitamin D₃ when the skin is exposed to ultraviolet sun rays.

In the liver mainly, Vitamin D₃ is metabolised into 25-Hydroxyvitamin D₃ (25OH D₃) which is the main form of Vitamin D circulating in the body.

25OH D₃ is a precursor for other Vitamin D metabolites and has also a limited activity by itself.

The most active derivative is 1,25-hydroxyvitamin D₃, produced in the kidney (or placenta) by 1-hydroxylation of 25OH D₃.

25OH Vitamin D stimulates the intestinal absorption of both calcium and phosphorus and also bone resorption and mineralisation.

25OH Vitamin D might also be active in other tissues responsible for calcium transport (placenta, kidney, mammary gland ...) and endocrine gland (parathyroid glands, beta cells...).


Vitamin D₃ and Vitamin D₂ are also available by ingestion through food or dietary supplementation.

IV. PRINCIPLES OF THE METHOD

The DIIAsource 25OH Vitamin D Total ELISA is a solid phase Enzyme Linked Immunosorbent Assay performed on microtiterplates. During a first 2 hours incubation step, at room temperature, total 25OH Vitamin D (D₂ and D₃) present in calibrators, controls and samples is dissociated from binding serum proteins to fix on binding sites of a specific monoclonal antibody. After 1 washing step, a fixed amount of 25OH Vitamin D-labelled with biotin in presence of horseradish peroxidase (HRP), compete with unlabelled 25OH Vitamin D₂ and 25OH Vitamin D₃ present on the binding sites of the specific monoclonal antibody. After a 30 minutes incubation at room temperature, the microtiterplate is washed to stop the competition reaction. The Chromogene solution (TMB) is added and incubated for 15 minutes. The reaction is stopped with the addition of Stop Solution and the microtiterplate is then read at the appropriate wavelength. The amount of substrate turnover is determined colourimetrically by measuring the absorbance, which is inversely proportional to the total 25OH Vitamin D (D₂ and D₃) concentration.

A calibration curve is plotted and the total 25OH Vitamin D (D₂ and D₃) concentrations of the samples are determined by dose interpolation from the calibration curve.

V. REAGENTS PROVIDED

Reagents	96 Tests Kit	Reconstitution
 Microtiterplate (96 breakable wells) with anti 25OH Vit. D2 and D3 (monoclonal antibodies)	96 wells	Ready for use
CAL 0 Calibrator 0: biological matrix with gentamycin and proclin	1 vial lyophilised	Add 1 ml distilled water
CAL N Calibrators 1-5 (see exact values on vial labels) in horse serum with gentamycin and proclin	5 vials lyophilised	Add 1 ml distilled water
CONTROL N Controls N = 2 in human serum with proclin	2 vials lyophilised	Add 1 ml distilled water
INC BUF Incubation Buffer with casein and proclin	1 vial 30 ml	Ready for use
CONJ CONC 25OH Vit D Concentrated Conjugate	1 vial 0.3 ml	Dilute 100 x with conjugate buffer
CONJ BUF Conjugate Buffer with casein and proclin	1 vial 30 ml	Ready for use
HRP CONC Concentrated HRP	1 vial 0.2 ml	Dilute 200 x with conjugate buffer
WASH SOLN CONC Wash solution (TRIS-HCl)	1 vial 10 ml	Dilute 200 x with distilled water (use a magnetic stirrer).
CHROM TMB Chromogenic solution TMB (Tetramethylbenzidine)	1 vial 13 ml	Ready for use
STOP SOLN Stop solution 0.2M H ₂ SO ₄	1 vial 13 ml	Ready for use

Note:

For dilution of samples having concentrations of 25OH Vitamin D above the highest calibrator concentration, use Control 1 or a serum sample with a concentration of 25OH below 25ng/mL, and above 4.4ng/mL (limit of quantification of the assay), as measured in this assay. Use Ctrl 1 or this sample to dilute 2X the out of curve samples. Take the concentration of the Ctrl 1* or the low sample into account when calculating the dilution result.

* Use the concentration of Ctrl 1 measured in the same run as the dilution run, not the mean concentration on the Ctrl 1 label!

Calculations:

Sample value = (Measured value – F1*Measured Ctrl 1) / F2
with the following values for F1 and F2:

- Sample diluted 2 times, F1 = 0.5; F2 = 0.5
- Sample diluted 4 times, F1 = 0.75; F2 = 0.25
- Sample diluted 8 times, F1 = 0.875; F2 = 0.125

Example:

A sample out of the calibration curve is diluted 4 times with Ctrl 1, and is measured at 70ng/mL. Ctrl 1 is measured in the same run at 20ng/mL. Dilution 4 times, F1 = 0.75; F2 = 0.25

Sample calculated value = (70 – 0.75*20) / 0.25 = 220ng/mL

No international reference material is available

VI. SUPPLIES NOT PROVIDED

The following material is required but not provided in the kit:

1. Distilled water
2. Pipettes for delivery of: 25 µl, 250 µl and 1 ml (the use of accurate pipettes with disposable plastic tips is recommended)
3. Vortex mixer
4. Magnetic stirrer
5. Plate shaker (400 rpm)
6. Washer for microtiterplates
7. Microtiterplate reader capable of reading at 450 nm and 650 nm (bichromatic reading)

VII. REAGENT PREPARATION

- Calibrator 0:** Reconstitute the Calibrator 0 with 1 ml distilled water.
- Calibrators 1-5:** Reconstitute the Calibrators 1-5 with 1 ml distilled water.
- Controls:** Reconstitute the Controls with 1 ml distilled water.
- Working HRP conjugate solution:**

! The working HRP conjugate solution is to be prepared during the incubation and minimum 1h45 minutes before its use (cf X.B.5).

Prepare an adequate volume of working HRP conjugate solution by mixing the 3 reagents in the following sequence: (1) Conjugate buffer, (2) Concentrated Conjugate, (3) Vortex, (4) Concentrated HRP, (5) Vortex.

The order of addition of those 3 reagents is critical and should be rigorously respected to get reproducible Optical Densities.

Prepare the solution according to the number of used strips, as indicated in the below table: example for 6 strips (48 wells): 130 µl of concentrated conjugate and 65 µl of concentrated HRP to 13 ml of conjugate buffer.

Use a vortex to homogenize.

Until its use, keep the working HRP conjugate at room temperature and avoid direct sunlight or use a brown glass vial for its preparation.

The preparation of working HRP conjugate is not stable and must be discarded if not used.

Nb.of strips	Volume of Conjugate Buffer (ml)	Volume of Concentrated Conjugate (µl)	Volume of Concentrated HRP (µl)
1	3	30	15
2	5	50	25
3	7	70	35
4	9	90	45
5	11	110	55
6	13	130	65
7	15	150	75
8	17	170	85
9	19	190	95
10	21	210	105
11	23	230	115
12	25	250	125

- Working Wash solution:** Prepare an adequate volume of Working Wash solution by adding 199 volumes of distilled water to 1 volume of Wash Solution (200x). Use a magnetic stirrer to homogenize. Discard unused Working Wash solution at the end of the day.

VIII. STORAGE AND EXPIRATION DATING OF REAGENTS

- Before opening or reconstitution, all kits components are stable until the expiry date, indicated on the label, if kept at 2 to 8°C.
- After reconstitution, calibrators and controls are stable for eight weeks at 2 to 8°C. For longer storage periods, aliquots should be made and kept at -20°C for maximum 4 months. Avoid subsequent freeze-thaw cycles.

- Freshly prepared Working Wash solution should be used on the same day.
- Alterations in physical appearance of kit reagents may indicate instability or deterioration.

IX. SAMPLE COLLECTION AND PREPARATION

This kit is suitable for serum and heparinized plasma samples. Serum and heparinized plasma samples must be kept at 2-8°C. If the test is not run within 24 hrs, **sampling and storage at -20°C is recommended.**

Avoid subsequent freeze-thaw cycles.

Serum and heparinized plasma provide similar results. $Y(\text{Heparin plasma}) = 0.9922 \times (\text{serum}) + 0.2129 \text{ ng/ml}$, $r^2 = 0.9944$, $n = 10$

PROCEDURE

A. Handling notes

- Do not use the kit or components beyond expiry date.
- Do not mix materials from different kit lots.
- Bring all the reagents to room temperature prior to use.
- Thoroughly mix all reagents and samples by gentle agitation or swirling.
- Perform calibrators, controls and samples in duplicate. Vertical alignment is recommended.
- Use a clean plastic container to prepare the Wash Solution.
- In order to avoid cross-contamination, use a clean disposable pipette tip for the addition of each reagent and sample.
- For the dispensing of the Chromogenic Solution and the Stop Solution avoid pipettes with metal parts.
- High precision pipettes or automated pipetting equipment will improve the precision.
- Respect the incubation times.
- To avoid drift, the time between pipetting of the first calibrator and the last sample must be limited to the time mentioned in section XIII paragraph E (Time delay).**
- Prepare a calibration curve for each run, do not use data from previous runs.
- Dispense the Chromogenic Solution within 15 minutes following the washing of the microtiterplate.
- During incubation with Chromogenic Solution, avoid direct sunlight on the microtiterplate.

B. Procedure

- Select the required number of strips for the run. The unused strips should be resealed in the bag with a desiccant and stored at 2-8°C.
- Secure the strips into the holding frame.
- Pipette 25 µl of each Calibrator, Control and Sample into the appropriate wells.
- Pipette 250 µl of Incubation Buffer into all the wells.
- Incubate for 2 hours at room temperature, on a plate shaker (400 rpm)
 - Prepare the Working HRP conjugate solution during the incubation and minimum 1h 45 minutes before its use.
- Aspirate the liquid from each well.
- Wash the plate 3 times by:
 - dispensing 0.35 ml of Wash Solution into each well
 - aspirating the content of each well
- Pipette 250 µl of the working HRP conjugate solution into each well. Incubate the microtiterplate for 30 minutes at room temperature, on a plate shaker (400 rpm).
- Aspirate the liquid from each well.
- Wash the plate 3 times by:
 - dispensing 0.35 ml of Wash Solution into each well
 - aspirating the content of each well
- Pipette 100 µl of the Chromogenic solution into each well within 15 minutes following the washing step.
- Incubate the microtiterplate for 15 minutes at room temperature, on a plate shaker (400 rpm), avoid direct sunlight.
- Pipette 100 µl of Stop Solution into each well.
- Read the absorbances at 450 nm (reference filter 630 nm or 650 nm) within 1 hour and calculate the results as described in section XI

XI. CALCULATION OF RESULTS

- Read the plate at 450 nm against a reference filter set at 650 nm (or 630 nm).
- Calculate the mean of duplicate determinations.

- We recommend the use of computer assisted methods to construct the calibration curve. 4-parameter logistic function curve fitting is the preferred method. Reject obvious outliers.
- By interpolation of the sample OD values, determine the 25OH Vitamin D concentrations of the samples from the calibration curve.

XII. TYPICAL DATA

The following data are for illustration only and should never be used instead of the real time calibration curve.

25OH-ELISA		OD units
Calibrator	0 ng/ml	2.79
	3.44 ng/ml	2.56
	16.28 ng/ml	1.93
	32.39 ng/ml	1.29
	63.54 ng/ml	0.49
	122.76 ng/ml	0.16

Note: 1 ng/ml = 2.5 pmol/ml

XIII. PERFORMANCE AND LIMITATIONS

A. Limits of Detection

The Limit of Blank (LoB), Limit of Detection (LoD), and the Limit of Quantitation (LoQ), were determined in accordance with the CLSI guideline EP17-A.

The LoB was calculated by measuring the blank several times and calculating the 95th percentile of the distribution of the test values. The LoB was calculated to be 2.07 ng/ml.

The LoD was calculated as described in the guideline. The LoD was calculated to be 5.26 ng/ml.

The LoQ was calculated by testing 5 samples of low value 10 times in different test. The LoQ was calculated to be 3.35ng/ml with CV of 20%.

B. Specificity

Cross reactivity of the 25OH Vitamin D Total ELISA assay was determined by testing sera with spiked and unspiked cross reactants. The results are summarized in the following table:

Compound and Concentration	% Cross reaction
25OH-Vitamin D ₃ at 10 ng/mL	92
25OH-Vitamin D ₂ at 10 ng/mL	91
1,25(OH) ₂ -Vitamin D ₃ at 200 ng/mL	3.10
1,25(OH) ₂ -Vitamin D ₂ at 667 ng/mL	0.35
Vitamin D ₃ at 200 ng/mL	0.17
3-epi-25OH-Vitamin D ₃ at 20 µg/mL	0.91

The effect of potential interfering substances on samples using the DAsoure 25 OH Vitamin D Total ELISA test was evaluated. Different levels of Hemoglobin, Triglyceride, Vitamin C, Bilirubin Conjugate and Unconjugated in serum samples were tested on samples with different 25OH Vitamin D Concentration. Our acceptance criteria was to have interference of less than 10%. The tested substances did not affect the performance of the DAsoure 25 OH Vitamin D Total ELISA test.

Substance	25OH Vitamin D (ng/mL)	Concentration of Interferent (mg/dL)	Mean % Variation
Hemoglobin	21.87	250	-3.7%
		500	
		750	
Bilirubin Conjugated	37.36	250	2.5%
		500	
		750	
Bilirubin Unconjugated	21.87	50	-0.2%
		100	
		200	
	37.36	50	
		100	
		200	

Triglyceride	21.87	6.25	-2.1%
		125	
		250	
	37.36	500	
		6.25	
		125	
Vitamin C	21.87	250	2.9%
		500	
		1	
	37.36	10	
		100	
		1	
Biotin	21.87	0.2	2.5%
		2	
		4	
	37.36	0.2	
		2	
		4	

C Precision

The assay precision was calculated by running samples for a span of at least 20 days on 3 different lots. The results are summarized in the table below:

INTRA-ASSAY				INTER-ASSAY			
Sample	N	<X> ± SD (ng/ml)	C.V. (%)	Sample	N	<X> ± SD (ng/ml)	C.V. (%)
A	24	20.40 ± 0.69	3.4	A	10	12.62 ± 0.79	6.3
B	24	33.32 ± 0.96	2.9	B	10	20.99 ± 0.69	3.3
				C	10	33.22 ± 1.30	3.9
				D	10	70.25 ± 2.42	3.4

SD: Standard Deviation, CV: Coefficient of variation

D. Reproducibility

The reproducibility of the assay was done by testing three samples in duplicate for five days, twice a day, at three sites with two technicians per site. The mean results are summarized in the table below:

Sample	n	ng/mL	Within-Run	Between-Run	Between-Day	Between-Tech	Between-Site	Total
1	57	25.5	SD 0.22	0.61	0.98	1.54	2.21	2.59
			CV 0.3%	0.9%	3.8%	6.0%	8.7%	10.2%
2	57	52.9	SD 0.64	1.57	1.11	2.28	4.79	5.19
			CV 0.9%	2.3%	2.1%	4.3%	8.1%	9.8%
3	59	124.9	SD 1.00	1.74	1.84	3.39	4.98	6.25
			CV 1.4%	2.5%	1.5%	2.7%	4.0%	5.0%

E. Accuracy

Recovery was assessed by adding different levels of 25OH Vitamin D to samples. The results are summarized in the table below:

RECOVERY TEST	
Added 25OH-Vit.D ₂ (ng/ml)	Recovery (%)
5	90
10	92
25	85
50	71
Added 25OH-Vit.D ₂ (ng/ml)	Recovery (%)
5	88
10	91
25	88
50	83

A sample with a concentration known to be distributed throughout the measurable range was tested at equidistant dilutions, according to the dilution protocol in chapter V, to determine the linear range of the assay. A linear regression analysis was performed. The results are summarized in the following table:

DILUTION TEST:			
Sample dilution	Theoretical concentration (ng/ml)	Measured concentration (ng/ml)	Recovery (%)
1/1	-	65.20	-
1/2	32.60	29.34	90.0
1/4	16.30	15.01	92.1
1/8	8.15	8.91	109.3

Samples were diluted with the zero calibrator.

The linear range of the assay was found to be 8.91 ng/mL to 65.20 ng/mL.

F. Time delay

Time delay test between the last Calibrator and sample dispensing results is shown in the following table.

TIME DELAY			
	0 min (ng/ml)	10 min (ng/ml)	20 min (ng/ml)
Sample 1	27.9	30.5	30.2
Sample 2	49.5	47.5	49.0


Assay results remain accurate even when incubation buffer is dispensed 10 and 20 minutes after the Calibrator has been added in the coated wells.

G. Limitations of the test

1. The test is an aid in the diagnosis and is to be used in conjunction with clinical findings.
2. The performance of this assay has not been established in a pediatric population.
3. Samples suspected of containing concentrations above the highest calibrator should be assayed in dilution.
4. Hemolysed samples should not be used.

H. Method comparison

The performance of the DIALsource 25OH Vitamin D Total ELISA test was determined by conducting a correlation study tested at three different sites using a total of 356 samples. The samples were tested on both the DIALsource 25OH Vitamin D Total ELISA test and a commercially available 25OH Vitamin D ELISA test. The results ranged from 8.0ng/ml to 123.0ng/ml, the correlation coefficient between the two methods was 0.917, with the 95% confidence interval of 87.6% to 93.6%, a slope of 0.954 and the y-intercept of 3.05. The following graph summarizes the results:



XIV. INTERNAL QUALITY CONTROL

- If the results obtained for Control 1 and/or Control 2 are not within the range specified on the vial label, the results cannot be used unless a satisfactory explanation for the discrepancy has been given.
- If desirable, each laboratory can make its own pools of control samples, which should be kept frozen in aliquots. Controls which contain azide will interfere with the enzymatic reaction and cannot be used.
- Acceptance criteria for the difference between the duplicate results of the samples should rely on Good Laboratory Practises
- It is recommended that Controls be routinely assayed as unknown samples to measure assay variability. The performance of the assay should be monitored with quality control charts of the controls.
- It is good practise to check visually the curve fit selected by the computer.

XV PRECAUTIONS AND WARNINGS

Safety

For research use only.

The human blood components included in this kit have been tested by European approved and/or FDA approved methods and found negative for HBsAg, anti-HCV, anti-HIV-1 and 2. No known method can offer complete assurance that human blood derivatives will not transmit hepatitis, AIDS or other infections. Therefore, handling of reagents, serum samples should be in accordance with local safety procedures.

All animal products and derivatives have been collected from healthy animals. Bovine components originate from countries where BSE has not been reported. Nevertheless, components containing animal substances should be treated as potentially infectious.

Avoid any skin contact with all reagents, Stop Solution contains H₂SO₄. In case of contact, wash thoroughly with water.

Do not smoke, drink, eat or apply cosmetics in the working area. Do not pipette by mouth. Use protective clothing and disposable gloves.

For more information, refer to the MSDS.

XVI. BIBLIOGRAPHY

- ZERWEKH J.E. (2008)
Blood biomarkers of Vitamin D status.
Am. J. Clin. Nutr., 87(suppl):1067S-91S.
- HOLICK M.F. (2006)
Resurrection of Vitamin D deficiency and rickets.
J. Clin. Invest., 116:2062-2072.
- HEANEY R.P. (2000)
Vitamin D: how much do we need, and how much is too much.
Osteoporos. Int., 11:553-555.
- DAWSON-HUGHES B., HEANEY R.P., HOLICK M.F., LIPS P., MEUNIER P.J. (1997)
Prevalence of Vitamin D insufficiency in an adult normal population.
Osteoporos. Int., 7:439-443.
- BISCHOFF-FERRARI H.A., GIOVANNUCCI E., WILLETT W.C., DIETRICH T., DAWSON-HUGHES B. (2006)
Estimation of optimal serum concentrations of 25-hydroxyvitamin D for multiple health outcomes.
Am. J. Clin. Nutr., 84:18-28.
- HOLICK M.F.(2004)
Sunlight and vitamin D for bone health and prevention of autoimmune

diseases, cancers and cardiovascular disease.

Am. J. Clin. Nutr., 80:16788S-1688S.

- HEANEY R.P. (2010)
Defining deficiency of vitamin D.
Clinical Laboratory International October 2010, vol.34: 16-19.
- HOLICK M.F. (2007)
Vitamin D deficiency.
N. Engl. J. Med., 357:266-281.
- TAHA N. M., VIETH R.(2010)
The problem of an optimal target level for 25-Hydroxyvitamin D, the test for vitamin D nutritional status.
Clinical Laboratory International, November 2010, vol.34: 28-30
- HOLICK M.F. (2009)
Vitamin D Status: Measurement, Interpretation, and Clinical Application
Ann. Epidemiol., 19:73-78.
- National Osteoporosis Foundation. Prevention – Vitamin D**
<http://www.nof.org/aboutosteoporosis/prevention/vitamingd>
- EP17-A
Protocols for Determination of Limits of Detection and Limits of Quantitation; Approved Guideline, STANDARD published by Clinical and Laboratory Standards Institute.

XVII. SUMMARY OF THE PROTOCOL

	CALIBRATORS (µl)	SAMPLE(S) CONTROLS (µl)
Calibrators (0-5)	25	-
Controls, Samples	-	25
Incubation Buffer	250	250
Incubate for 2 hours at room temperature with continuous shaking at 400 rpm. Prepare the working HRP conjugate during the incubation and minimum 1h 45 minutes before its use. The sequence of preparation is critical, see VII. Reagent Preparation		
Aspirate the contents of each well. Wash 3 times with 350 µl of Wash Solution and aspirate.		
Working HRP Conjugate	250	250
Incubate for 30 minutes at room temperature with continuous shaking at 400 rpm. Aspirate the contents of each well. Wash 3 times with 350 µl of Wash Solution and aspirate.		
Chromogenic Solution	100	100
Incubate for 15 min at room temperature with continuous shaking at 400 rpm.		
Stop Solution	100	100
Read on a microtiterplate reader Record the absorbance of each well at 450 nm (versus 630 or 650 nm).		

DIAsource Catalogue
Nr : KAR1971

Revision nr :
220516

Revision date : 16/05/2022